

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-003060

Location: CARR (Foster G.S.O. CARRINGTON P.D.)

- Check When Done
- A. Is the breath tube heated? ✓
  - B. Is the diagnostic check complete? ✓
  - C. Is the time, date, and year correct (re-set if necessary)? ✓
  - D. Print test. (Attach test record.) ✓
  - E. Low AC. Use  $\leq 0.03$  AC in ACA mode. (Attach test record.)  
Sim SN DR5143 Lot No. 201501B Rep AC 0.020 ✓
  - F. High AC. Use  $\geq 0.25$  AC in ACA mode. (Attach test record.)  
Sim SN DR3455 Lot No. 14270 Rep AC 0.300 ✓
  - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.  
(Attach test record.)  
Sim SN DR5132 Lot No. ICS #2 Solution 0.10AC + 0.05% ACE ✓
  - H. RFI check. Run CMS mode. Key radio during CMS test.  
(Attach test record.) ✓
  - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.  
(Attach test record.) 0.080 AC

Lot No. <u>33913080A4</u>	Cyl No. <u>18</u>	Exp. Date <u>1/1/2016</u>
Test 1 <u>0.079</u>	Test 4 <u>0.079</u>	Test 7 <u>0.079</u>
Test 2 <u>0.079</u>	Test 5 <u>0.079</u>	Test 8 <u>0.078</u>
Test 3 <u>0.080</u>	Test 6 <u>0.079</u>	Test 9 <u>0.079</u>
Average <u>0.079</u>		

J. Tank pressure, Level 3 Func D Sub G P: 982 psi Reg 1000 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

5.4.15  
Date

04 May 15  
Date

Charles E. Egan  
Field Inspector's Signature

Deb Kashner  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003060  
Location = CARR                    8164.13.00 06/09  
05/04/2015                            11:43

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer (Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            33913080A4  
Standard Cyl #?            18  
Standard Expiration?        01/01/2016  
Oper No?                    666666

Flow Cal. Date:            04/08/2008  
Slope                      702  
Intercept                  -647610

IR Calibration Date:        06/04/2008  
                                  3um                            9um

-----  
0th Coef(\*100):            -5190                        -7737  
1st Coef(\*100):            270580                      138779  
2nd Coef(\*100):            2499                         956  
H2O adj(mg/l\*10k):        377                          188

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003060  
Location = CARR      8164.13.00 06/09  
05/04/2015      12:09

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:10
02 Std. Sol.	0.019	12:11
03 Room Air	0.000	12:11
04 Std. Sol.	0.020	12:12
05 Room Air	0.000	12:12
06 Std. Sol.	0.020	12:13
07 Room Air	0.000	12:14


08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 201501B

County = 16

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

Low AC  
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003060  
Location = CARR      8164.13.00 06/09  
05/04/2015      12:15

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:16
02 Std. Sol.	0.289	12:16
03 Room Air	0.000	12:17
04 Std. Sol.	0.295	12:18
05 Room Air	0.000	12:18
06 Std. Sol.	0.296	12:19
07 Room Air	0.000	12:20

08 Sim Temp = 34.0°C

Simul Ser No = DR3455

Std Sol No = 14270

County = 16

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

HIGH AC  
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003060  
Location = CARR                    8164.13.00 06/09  
05/04/2015                            11:58

Test	AC	Time
01 Room Air	0.000	11:59
02 *Subject Test	INT*	12:00
03 Room Air	0.000	12:01

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 33913080A4  
Cyl No = 18  
Expiration Date = 01/01/2016  
County = 16                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Interferent Check*  
*0.10AC + 0.05% ACETONE*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003060  
Location = CARR            8164.13.00 06/09  
05/04/2015                            12:02

Test	AC	Time
01 Diagnostic	OK	12:03
02 Room Air	0.000	12:03
03 *Subject Test	RFI*	12:03
04 Room Air	0.000	12:04

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                            Weight = 150  
Test = DUI                                    Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 33913080A4  
Cyl No = 18  
Expiration Date = 01/01/2016  
County = 16                                    Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

*RFI CHECK*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003060  
Location = CARR      8164.13.00 06/09  
05/04/2015      11:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:55
02 Std. Gas	0.079	11:55
03 Room Air	0.000	11:56
04 Std. Gas	0.079	11:56
05 Room Air	0.000	11:57
06 Std. Gas	0.080	11:57
07 Room Air	0.000	11:58

Lot No = 33913080A4  
Cyl No = 18  
Exp Date = 01/01/2016  
County = 16      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003060  
Location = CARR      8164.13.00 06/09  
05/04/2015      12:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:05
02 Std. Gas	0.079	12:05
03 Room Air	0.000	12:06
04 Std. Gas	0.079	12:06
05 Room Air	0.000	12:07
06 Std. Gas	0.079	12:07
07 Room Air	0.000	12:08

Lot No = 33913080A4  
Cyl No = 18  
Exp Date = 01/01/2016  
County = 16      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*CALIBRATION CHECK*  
*0.080 AC*

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003060  
Location = CARR      8164.13.00 06/09  
05/04/2015      12:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:23
02 Std. Gas	0.079	12:24
03 Room Air	0.000	12:24
04 Std. Gas	0.078	12:24
05 Room Air	0.000	12:25
06 Std. Gas	0.079	12:25
07 Room Air	0.000	12:26


Lot No = 33913080A4

Cyl No = 18

Exp Date = 01/01/2016

County = 16

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080 AC

Form 106-I8000