



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59281 (03-2014)

Serial Number <i>80-005947</i>	Instrument Location <i>BSPD</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): *None*
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Pat Reng</i>	Date <i>5-4-15</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date <i>04 May 15</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = BSPD 8164.13.00 06/09
05/04/2015 15:29

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 29614080A2
Standard Cyl #? 27
Standard Expiration? 12/05/2016
Oper No? 029710

Flow Cal. Date: 06/21/2013
Slope 693
Intercept -735360.

IR Calibration Date: 04/06/2015
3um 9um

0th Coef(*100): -28096 -26982
1st Coef(*100): 280034 136042
2nd Coef(*100): 3530 1567
H2O adj(mg/l*10k): 510 437

***** Printer Test End *****


Operator Signature
PATRICK RENZ

Remarks: *Install new Intoxilyzer*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = BSPD 8164.13.00 06/09
05/04/2015 15:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:30
02 Std. Gas	0.080	15:31
03 Room Air	0.000	15:31
04 Std. Gas	0.079	15:31
05 Room Air	0.000	15:32
06 Std. Gas	0.080	15:32
07 Room Air	0.000	15:33

Lot No = 29614080A2
Cyl No = 27
Exp Date = 12/05/2016
County = 08 Oper No. = 029710



Operator Signature
PATRICK RENZ

Remarks: *Install new Intoxilyzer*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = BSPD 8164.13.00 06/09
05/04/2015 15:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	15:36
02 Room Air	0.000	15:36

*Invalid Test
Inhibited - RFI

Lot No = 29614080A2
Cyl No = 27
Exp Date = 12/05/2016
County = 08 Oper No. = 029710



Operator Signature
PATRICK RENZ

Remarks: *Install new Intoxilyzer*