



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 60002 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>William Clark</u>		
Location: <u>Bldg 547 / Innot A/B ND</u>		
Intoxilyzer® Serial Number: <u>80-004936</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>49</u>	Expiration Date: <u>9-3-2011</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: <u>0.076</u> % Test 2: <u>0.077</u> % Test 3: <u>0.077</u> %		
(ATTACH TEST RECORD)		

William Clark
Chemical Test Operator Signature

5-5-2010
Date

Deb Kashner
Reviewed By

12 May 2010
Date

80-004936
p. 33/34 PM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = MAFB 8164.13.00 06/09
05/05/2010 07:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	07:20
02 Std. Gas	0.076	07:21
03 Room Air	0.000	07:21
04 Std. Gas	0.077	07:21
05 Room Air	0.000	07:22
06 Std. Gas	0.077	07:22
07 Room Air	0.000	07:23

Lot No = 659358
Cyl No = 49
Exp Date = 09/03/2011
County = 51 Oper No. = 020809



Operator Signature
WILLIAM CLARK

Remarks:

Form 106-I8000

80-004936
P. 34/34 194