

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-003060

Location: CARR

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5115 Lot No. 12111 Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5143 Lot No. 13080 Rep AC 0.360 ✓
 - G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3375 Lot No. 1652 Rep AC 0.10 Ac + 0.05% acetone ✓
 - H. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. <u>16512080A2</u>	Cyl No. <u>11</u>	Exp. Date <u>08/01/2014</u>
Test 1 <u>0.078</u>	Test 4 <u>0.078</u>	Test 7 <u>0.078</u>
Test 2 <u>0.079</u>	Test 5 <u>0.079</u>	Test 8 <u>0.078</u>
Test 3 <u>0.079</u>	Test 6 <u>0.079</u>	Test 9 <u>0.078</u>
Average <u>0.078</u>		

J. Tank pressure, Level 3 Func D Sub G P: 952 psi Reg 950 psi

K. Remarks/Maintenance record: Annual inspection

05 May 2014
Date

05 MAY 2014
Date

Deb Kashner
Field Inspector's Signature

Charles E. Egan
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003060
Location = CARR 8164.13.00 06/09
05/05/2014 13:35

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:36
02 Std. Sol.	0.019	13:37
03 Room Air	0.000	13:37
04 Std. Sol.	0.019	13:38
05 Room Air	0.000	13:38
06 Std. Sol.	0.019	13:39
07 Room Air	0.000	13:40

08 Sim Temp = 34.0°C

Simul Ser No = DR5115

Std Sol No = 12111

County = 16

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

low AC 0.000 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003060
Location = CARR 8164.13.00 06/09
05/05/2014 13:45

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:45
02 Std. Sol.	0.289	13:46
03 Room Air	0.000	13:47
04 Std. Sol.	0.290	13:47
05 Room Air	0.000	13:48
06 Std. Sol.	0.293	13:49
07 Room Air	0.000	13:49


08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 13080

County = 16

Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

High AC ^{DK} ~~0.0~~ 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003060
Location = CARR 8164.13.00 06/09
05/05/2014 13:56

Test	AC	Time
01 Diagnostic	OK	13:57
02 Room Air	RFI*	13:57
03 Room Air	0.000	13:57

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 11
Expiration Date = 08/01/2014
County = 16 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:

RFI check
Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003060
Location = CARR 8164.13.00 06/09
05/05/2014 13:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:23
02 Std. Gas	0.078	13:24
03 Room Air	0.000	13:24
04 Std. Gas	0.079	13:24
05 Room Air	0.000	13:25
06 Std. Gas	0.079	13:25
07 Room Air	0.000	13:26

Lot No = 16512080A2
Cyl No = 11
Exp Date = 08/01/2014
County = 16 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC
Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003060
Location = CARR 8164.13.00 06/09
05/05/2014 13:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:31
02 Std. Gas	0.078	13:31
03 Room Air	0.000	13:32
04 Std. Gas	0.078	13:32
05 Room Air	0.000	13:32
06 Std. Gas	0.078	13:33
07 Room Air	0.000	13:33

Lot No = 16512080A2
Cyl No = 11
Exp Date = 08/01/2014
County = 16 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.080 AC