

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004946

Location: LINT

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5115 Lot No. 12111 Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5143 Lot No. 13080 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3375 Lot No. ICS#2 Rep AC 0.100 AC + 0.05% ACETONE ✓
 - H. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 16512080A2 Cyl No. 016 Exp. Date 08/01/2014

Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.080</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>

Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: ^{1005 CEE}~~1000~~ psi Reg 1000 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

02 MAY 14
Date

Charles E. Edin
Field Inspector's Signature

05 May 14 ✓
Date

Deb Kashner
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = LINT 8164.13.00 06/09
05/02/2014 08:45

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 16512080A2
Standard Cyl #? 016
Standard Expiration? 08/01/2014
Oper No? 666666

Flow Cal. Date: 01/18/2010
Slope 717
Intercept -728279

IR Calibration Date: 05/19/2011
 3um 9um

0th Coef(*100): -2231 -4869
1st Coef(*100): 282398 143476
2nd Coef(*100): 146 780
H2O adj(mg/l*10k): 303 375

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = LINT 8164.13.00 06/09
05/02/2014 09:07

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:08
02 Std. Sol.	0.022	09:08
03 Room Air	0.000	09:09
04 Std. Sol.	0.023	09:10
05 Room Air	0.000	09:10
06 Std. Sol.	0.024	09:11
07 Room Air	0.000	09:11

08 Sim Temp = 34.0°C

Simul Ser No = DR5115

Std Sol No = 12111

County = 15

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC

0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = LINT 8164.13.00 06/09
05/02/2014 09:12

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:13
02 Std. Sol.	0.296	09:14
03 Room Air	0.000	09:14
04 Std. Sol.	0.299	09:15
05 Room Air	0.000	09:15
06 Std. Sol.	0.301	09:16
07 Room Air	0.000	09:17

08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 13080

County = 15

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = LINT 8164.13.00 06/09
05/02/2014 09:02

Test	AC	Time
01 Room Air	0.000	09:03
02 *Subject Test	INT*	09:04
03 Room Air	0.000	09:04

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 16512080A2
Cyl No = 016
Expiration Date = 08/01/2014
County = 15 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT CHECK*
0.100AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = LINT 8164.13.00 06/09
05/02/2014 08:54

Test	AC	Time
01 Diagnostic	OK	08:55
02 Room Air	RFI*	08:55
03 Room Air	0.000	08:56

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 16512080A2
Cyl No = 016
Expiration Date = 08/01/2014
County = 15 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = LINT 8164.13.00 06/09
05/02/2014 08:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:58
02 Std. Gas	0.081	08:58
03 Room Air	0.000	08:59
04 Std. Gas	0.081	08:59
05 Room Air	0.000	08:59
06 Std. Gas	0.080	09:00
07 Room Air	0.000	09:00

Lot No = 16512080A2
Cyl No = 016
Exp Date = 08/01/2014
County = 15 Oper No. = 666666


Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = LINT 8164.13.00 06/09
05/02/2014 09:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:19
02 Std. Gas	0.081	09:19
03 Room Air	0.000	09:20
04 Std. Gas	0.081	09:20
05 Room Air	0.000	09:21
06 Std. Gas	0.081	09:21
07 Room Air	0.000	09:21

Lot No = 16512080A2
Cyl No = 16
Exp Date = 08/01/2014
County = 15 Oper No. = 666666


Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = LINT 8164.13.00 06/09
05/02/2014 09:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:22
02 Std. Gas	0.081	09:23
03 Room Air	0.000	09:23
04 Std. Gas	0.081	09:24
05 Room Air	0.000	09:24
06 Std. Gas	0.081	09:24
07 Room Air	0.000	09:25

Lot No = 16512080A2
Cyl No = 16
Exp Date = 08/01/2014
County = 15 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC