

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005360

Location: ASHL

- Check When Done
- A. Is the breath tube heated? ✓
  - B. Is the diagnostic check complete? ✓
  - C. Is the time, date, and year correct (re-set if necessary)? ✓
  - D. Print test. (Attach test record.) ✓
  - E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)  
Sim SN DR5115 Lot No. 12111 Rep AC 0.020 ✓
  - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)  
Sim SN DR5143 Lot No. 13080 Rep AC 0.300 ✓
  - G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.  
(Attach test record.)  
Sim SN DR3375 Lot No. ICS#2 Rep AC 0.100 AC + 0.05% ACETONE ✓
  - H. RFI check. Run CMS mode. Key radio on first room air.  
(Attach test record.) ✓
  - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.  
(Attach test record.)

Lot No. 33913080A4 Cyl No. 041 Exp. Date 01/01/2016

Test 1 <u>0.083</u>	Test 4 <u>0.082</u>	Test 7 <u>0.083</u>
Test 2 <u>0.083</u>	Test 5 <u>0.083</u>	Test 8 <u>0.082</u>
Test 3 <u>0.083</u>	Test 6 <u>0.081</u>	Test 9 <u>0.082</u>

Average 0.082

J. Tank pressure, Level 3 Func D Sub G P: 1015 psi Reg 1000+ psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

02 MAY 14  
Date

Charles E. Edler  
Field Inspector's Signature

05 May 14 ✓  
Date

Deb Kashner  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005360  
Location = ASHL      8164.13.00 06/09  
05/02/2014      10:36

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup


Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On Memfull? Yes  
# of Print Copies? 1  
Select Std (D/W/I)? Dry  
Standard Value? 0.080  
Standard Lot #? 33913080A4  
Standard Cyl #? 41  
Standard Expiration? 01/01/2016  
Oper No? 666666

Flow Cal. Date: 08/17/2011  
Slope 639  
Intercept -343114

IR Calibration Date: 08/17/2011  
3um 9um

0th Coef(*100):	-17266	-25747
1st Coef(*100):	263773	142804
2nd Coef(*100):	2606	1406
H2O adj(mg/l*10k):	499	577

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005360  
Location = ASHL                    8164.13.00 06/09  
05/02/2014                            11:01

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:01
02 Std. Sol.	0.018	11:02
03 Room Air	0.000	11:03
04 Std. Sol.	0.019	11:03
05 Room Air	0.000	11:04
06 Std. Sol.	0.018	11:05
07 Room Air	0.000	11:05

08 Sim Temp = 34.0°C

Simul Ser No = DR5115

Std Sol No = 12111

County = 26

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Low AC*

*0.020 AC*

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005360  
Location = ASHL      8164.13.00 06/09  
05/02/2014      10:58

Test	AC	Time
01 Room Air	0.000	10:59
02 *Subject Test	INT*	10:59
03 Room Air	0.000	11:00

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 33913080A4  
Cyl No = 41  
Expiration Date = 01/01/2016  
County = 26      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*INTERFERENT CHECK*  
*0.100AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005360  
Location = ASHL                    8164.13.00 06/09  
05/02/2014                            10:55

Test	AC	Time
01 Diagnostic	OK	10:56
02 Room Air	RFI*	10:56
03 Room Air	0.000	10:56

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 33913080A4  
Cyl No = 41  
Expiration Date = 01/01/2016  
County = 26                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:    *RFI CHECK*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005360  
Location = ASHL      8164.13.00 06/09  
05/02/2014      11:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:13
02 Std. Gas	0.083	11:13
03 Room Air	0.000	11:13
04 Std. Gas	0.083	11:14
05 Room Air	0.000	11:14
06 Std. Gas	0.083	11:15
07 Room Air	0.000	11:15

Lot No = 33913080A4  
Cyl No = 41  
Exp Date = 01/01/2016  
County = 26      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005360  
Location = ASHL      8164.13.00 06/09  
05/02/2014      11:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:16
02 Std. Gas	0.082	11:16
03 Room Air	0.000	11:17
04 Std. Gas	0.083	11:17
05 Room Air	0.000	11:18
06 Std. Gas	0.081	11:18
07 Room Air	0.000	11:19

Lot No = 33913080A4

Cyl No = 41

Exp Date = 01/01/2016

County = 26

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK

0.080 AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005360  
Location = ASHL      8164.13.00 06/09  
05/02/2014      11:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:19
02 Std. Gas	0.083	11:20
03 Room Air	0.000	11:20
04 Std. Gas	0.082	11:21
05 Room Air	0.000	11:21
06 Std. Gas	0.082	11:22
07 Room Air	0.000	11:22

Lot No = 33913080A4  
Cyl No = 41  
Exp Date = 01/01/2016  
County = 26      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080 AC

Form 106-I8000