

BrW-004

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004942

Location: BURK

BURKE
COUNTY,
S.O.

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5191 Lot No. 15160 Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3847 Lot No. 16060 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5191 Lot No. ICS4 Solution 0.10AC + 0.05% ✓
 - H. RFI check. Run CMS mode. Key radio during CMS test. ACETONE
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 13314080A2 Cyl No. 058 Exp. Date 6/5/2016

Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.081</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>

Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 1005 psi Reg ~1000 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION.

5/4/2016
Date

Charles Elder
Field Inspector's Signature

05 May 16
Date

Deb Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = BURK 8164.13.00 06/09
05/04/2016 10:56

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 13314080A2
Standard Cyl #? 58
Standard Expiration? 06/05/2016
Oper No? 666666

Flow Cal. Date: 08/21/2012
Slope 638
Intercept -487461

IR Calibration Date: 05/12/2015

	3um	9um
0th Coef(*100):	-11858	-18295
1st Coef(*100):	255387	131150
2nd Coef(*100):	3253	1510
H2O adj(mg/l*10k):	493	483

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks: **PRINT TEST**

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = BURK 8164.13.00 06/09
05/04/2016 11:09

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:10
02 Std. Sol.	0.018	11:11
03 Room Air	0.000	11:11
04 Std. Sol.	0.018	11:12
05 Room Air	0.000	11:12
06 Std. Sol.	0.019	11:13
07 Room Air	0.000	11:14


08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 15160

County = 07

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = BURK 8164.13.00 06/09
05/04/2016 11:15

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:15
02 Std. Sol.	0.309	11:25
03 Room Air	0.000	11:26
04 Std. Sol.	0.309	11:26
05 Room Air	0.000	11:27
06 Std. Sol.	0.310	11:28
07 Room Air	0.000	11:28

08 Sim Temp = 34.0°C

Simul Ser No = DR3847
Std Sol No = 16060
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = BURK 8164.13.00 06/09
05/04/2016 11:29

Test	AC	Time
01 Room Air	0.000	11:31
02 *Subject Test	INT*	11:31
03 Room Air	0.000	11:32

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 13314080A2
Cyl No = 58
Expiration Date = 06/05/2016
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = BURK 8164.13.00 06/09
05/04/2016 11:32

Test	AC	Time
01 Diagnostic	OK	11:33
02 Room Air	RFI*	11:33
03 Room Air	0.000	11:34

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 13314080A2
Cyl No = 58
Expiration Date = 06/05/2016
County = 07 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = BURK 8164.13.00 06/09
05/04/2016 11:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:35
02 Std. Gas	0.081	11:35
03 Room Air	0.000	11:36
04 Std. Gas	0.081	11:36
05 Room Air	0.000	11:37
06 Std. Gas	0.081	11:37
07 Room Air	0.000	11:37

Lot No = 13314080A2
Cyl No = 58
Exp Date = 06/05/2016
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = BURK 8164.13.00 06/09
05/04/2016 11:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:38
02 Std. Gas	0.081	11:38
03 Room Air	0.000	11:39
04 Std. Gas	0.081	11:39
05 Room Air	0.000	11:40
06 Std. Gas	0.081	11:40
07 Room Air	0.000	11:41

Lot No = 13314080A2
Cyl No = 58
Exp Date = 06/05/2016
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = BURK 8164.13.00 06/09
05/04/2016 11:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:41
02 Std. Gas	0.081	11:42
03 Room Air	0.000	11:42
04 Std. Gas	0.081	11:43
05 Room Air	0.000	11:43
06 Std. Gas	0.081	11:44
07 Room Air	0.000	11:44

Lot No = 13314080A2
Cyl No = 58
Exp Date = 06/05/2016
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC