

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005938

Location: STAN

(MOUNTAIN
COUNTY
S.O.)

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN ~~DR3803~~ Lot No. 15160 Rep AC 0.020
DR5191 CEE ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3847 Lot No. 16060 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5190 Lot No. ICS4 Solution 0.10 AC + 0.05% ACETONE ✓
 - H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 13314080A2 Cyl No. 054 Exp. Date 6/5/2016

Test 1 <u>0.079</u>	Test 4 <u>0.080</u>	Test 7 <u>0.079</u>
Test 2 <u>0.079</u>	Test 5 <u>0.080</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.079</u>	Test 9 <u>0.080</u>

Average 0.080

J. Tank pressure, Level 3 Func D Sub G P: 990 psi Reg ~1000 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION, REPLACED BREATH HOSE O-RING,

5/3/2016
Date

05 May 16
Date

Charles E. Edwards
Field Inspector's Signature

Deb Shearer
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005938
Location = STAN 8164.13.00 06/09
05/03/2016 12:32

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 13314080A2
Standard Cyl #? 054
Standard Expiration? 06/05/2016
Oper No? 666666

Flow Cal. Date: 07/15/2013
Slope 664
Intercept -578364

IR Calibration Date: 07/27/2015
 3um 9um

0th Coef(*100): -14232 -19463
1st Coef(*100): 289698 139293
2nd Coef(*100): 2380 1071
H2O adj(mg/l*10k): 253 294

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks: **PRINT TEST**

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005938
Location = STAN 8164.13.00 06/09
05/03/2016 13:01

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:02
02 Std. Sol.	0.018	13:02
03 Room Air	0.000	13:03
04 Std. Sol.	0.018	13:04
05 Room Air	0.000	13:04
06 Std. Sol.	0.018	13:05
07 Room Air	0.000	13:05

08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 15160

County = 31

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005938
Location = STAN 8164.13.00 06/09
05/03/2016 12:51

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:52
02 Std. Sol.	0.300	12:52
03 Room Air	0.000	12:53
04 Std. Sol.	0.301	12:54
05 Room Air	0.000	12:54
06 Std. Sol.	0.301	12:55
07 Room Air	0.000	12:55

08 Sim Temp = 34.0°C

Simul Ser No = DR3847

Std Sol No = 16060

County = 31

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC

0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005938
Location = STAN 8164.13.00 06/09
05/03/2016 12:40

Test	AC	Time
01 Room Air	0.000	12:41
02 *Subject Test	INT*	12:42
03 Room Air	0.000	12:42

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 13314080A2
Cyl No = 054
Expiration Date = 06/05/2016
County = 31 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005938
Location = STAN 8164.13.00 06/09
05/03/2016 12:43

Test	AC	Time
01 Diagnostic	OK	12:43
02 Room Air	0.000	12:44
03 *Subject Test	RFI*	12:44
04 Room Air	0.000	12:45

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 13314080A2
Cyl No = 054
Expiration Date = 06/05/2016
County = 31 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005938
Location = STAN 8164.13.00 06/09
05/03/2016 12:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:37
02 Std. Gas	0.079	12:38
03 Room Air	0.000	12:38
04 Std. Gas	0.079	12:39
05 Room Air	0.000	12:39
06 Std. Gas	0.080	12:40
07 Room Air	0.000	12:40

Lot No = 13314080A2
Cyl No = 054
Exp Date = 06/05/2016
County = 31 Oper No. = 666666


Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005938
Location = STAN 8164.13.00 06/09
05/03/2016 13:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:09
02 Std. Gas	0.080	13:09
03 Room Air	0.000	13:10
04 Std. Gas	0.080	13:10
05 Room Air	0.000	13:10
06 Std. Gas	0.079	13:11
07 Room Air	0.000	13:11

Lot No = 13314080A2
Cyl No = 054
Exp Date = 06/05/2016
County = 31 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: **CALIBRATION CHECK**
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005938
Location = STAN 8164.13.00 06/09
05/03/2016 13:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:12
02 Std. Gas	0.079	13:12
03 Room Air	0.000	13:13
04 Std. Gas	0.080	13:13
05 Room Air	0.000	13:14
06 Std. Gas	0.080	13:14
07 Room Air	0.000	13:15

Lot No = 13314080A2

Cyl No = 054

Exp Date = 06/05/2016

County = 31

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

CALIBRATION CHECK

0.080 AC

Form 106-I8000