

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005946

Location: WILL (WILLIAMS COUNTY CORR.)

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR3803 Lot No. 15160 Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3847 Lot No. 16060 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5190 Lot No. ICS4 Solution 0.10AC+0.05% ACETONE ✓
 - H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 29614080A2 Cyl No. 009 Exp. Date 12/05/2016

Test 1 <u>0.081 CEC</u>	Test 4 <u>0.081</u>	Test 7 <u>0.082</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.080</u>	Test 6 <u>0.082</u>	Test 9 <u>0.081</u>

Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 491 psi Reg ~500 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

5/2/2016
Date
05 May 16
Date

Charles E. Egan
Field Inspector's Signature
Deb Shanover
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = WILL 8164.13.00 06/09
05/02/2016 15:42

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 29614080A2
Standard Cyl #? 9
Standard Expiration? 12/05/2016
Oper No? 666666

Flow Cal. Date: 08/10/2015
Slope 673
Intercept -638348

IR Calibration Date: 08/10/2015

	3um	9um
0th Coef(*100):	-22507	-13665
1st Coef(*100):	273651	130615
2nd Coef(*100):	5249	2073
H2O adj(mg/l*10k):	617	504

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = WILL 8164.13.00 06/09
05/02/2016 15:50

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:51
02 Std. Sol.	0.019	15:52
03 Room Air	0.000	15:52
04 Std. Sol.	0.019	15:53
05 Room Air	0.000	15:53
06 Std. Sol.	0.020	15:54
07 Room Air	0.000	15:55

08 Sim Temp = 34.0°C

Simul Ser No = DR3803

Std Sol No = 15160

County = 53

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = WILL 8164.13.00 06/09
05/02/2016 15:55

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:56
02 Std. Sol.	0.300	15:57
03 Room Air	0.000	15:57
04 Std. Sol.	0.304	15:58
05 Room Air	0.000	15:59
06 Std. Sol.	0.306	15:59
07 Room Air	0.000	16:00

08 Sim Temp = 34.0°C

Simul Ser No = DR3847
Std Sol No = 16060
County = 53 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: HIGH AC
0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = WILL 8164.13.00 06/09
05/02/2016 16:02

Test	AC	Time
01 Room Air	0.000	16:03
02 *Subject Test	INT*	16:03
03 Room Air	0.000	16:04

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 9
Expiration Date = 12/05/2016
County = 53 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = WILL 8164.13.00 06/09
05/02/2016 16:05

Test	AC	Time
01 Diagnostic	OK	16:06
02 Room Air	0.000	16:07
03 *Subject Test	RFI*	16:07
04 Room Air	0.000	16:07

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 9
Expiration Date = 12/05/2016
County = 53 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = WILL 8164.13.00 06/09
05/02/2016 16:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:08
02 Std. Gas	0.081	16:08
03 Room Air	0.000	16:09
04 Std. Gas	0.081	16:09
05 Room Air	0.000	16:10
06 Std. Gas	0.080	16:10
07 Room Air	0.000	16:11

Lot No = 29614080A2
Cyl No = 9
Exp Date = 12/05/2016
County = 53 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = WILL 8164.13.00 06/09
05/02/2016 16:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:11
02 Std. Gas	0.081	16:12
03 Room Air	0.000	16:12
04 Std. Gas	0.081	16:12
05 Room Air	0.000	16:13
06 Std. Gas	0.082	16:13
07 Room Air	0.000	16:14

Lot No = 29614080A2
Cyl No = 9
Exp Date = 12/05/2016
County = 53 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC
Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = WILL 8164.13.00 06/09
05/02/2016 16:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:15
02 Std. Gas	0.082	16:15
03 Room Air	0.000	16:16
04 Std. Gas	0.081	16:16
05 Room Air	0.000	16:16
06 Std. Gas	0.081	16:17
07 Room Air	0.000	16:17

Lot No = 29614080A2
Cyl No = 9
Exp Date = 12/05/2016
County = 53 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC