

**INTOXILYZER 8000 INSPECTION**

Serial No.: 80-004936

Location: MAFB

- |  | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes?   | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?  | <u>✓</u>        |
| C. Is the breath tube heated?  | <u>✓</u>        |
| D. Is the diagnostic check complete?   | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?  | <u>✓</u>        |
| F. Print test. (Attach test record.)   | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5132</u> Lot No. <u>    </u> Rep AC <u>0.021</u>   | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5115</u> Lot No. <u>    </u> Rep AC <u>0.298</u>                                | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR5186</u> Lot No. <u>    </u> Rep AC <u>0.100</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)  | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)  |                 |

Lot No. <u>659358</u>	Cyl No. <u>49</u>	Exp. Date <u>09/03/2011</u>
Test 1 <u>0.076</u>	Test 4 <u>0.078</u>	Test 7 <u>0.078</u>
Test 2 <u>0.077</u>	Test 5 <u>0.078</u>	Test 8 <u>0.078</u>
Test 3 <u>0.077</u>	Test 6 <u>0.078</u>	Test 9 <u>0.078</u>
Average <u>0.078</u>		

L. Tank pressure, Level 3 Func D Sub G P: 790 psi Reg 800 psi

M. Remarks/Maintenance record: Annual inspection

<u>06 May 2010</u> Date	<u>Deb Kashner</u> Field Inspector's Signature
<u>15 May 2010</u> Date	<u>Mary Pearson</u> Reviewed by



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004936  
Location = MAFB      8164.13.00 06/09  
05/06/2010      11:18

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:19
02 Std. Sol.	0.016	11:19
03 Room Air	0.000	11:20
04 Std. Sol.	0.016	11:21
05 Room Air	0.000	11:21
06 Std. Sol.	0.017	11:22
07 Room Air	0.000	11:22

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 64

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Low AC 0.021 AC*

Form 106-18000

80-004936  
P. 43/49 KA

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004936  
Location = MAFB      8164.13.00 06/09  
05/06/2010      11:29

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:30
02 Std. Sol.	0.279	11:31
03 Room Air	0.000	11:31
04 Std. Sol.	0.279	11:32
05 Room Air	0.000	11:32
06 Std. Sol.	0.278	11:33
07 Room Air	0.000	11:34

08 Sim Temp = 34.0°C

Simul Ser No = 00000  
Std Sol No = 000  
County = 64      Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

*Linearity test 0.298 AC*

Form 106-18000

80-004936  
P. 44/49 JRM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004936  
Location = MAFB      8164.13.00 06/09  
05/06/2010      11:34

Test	AC	Time
01 Room Air	0.000	11:35
02 *Subject Test	INT*	11:35
03 Room Air	0.000	11:36

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Simul Ser No = 00000  
Std Sol No = 000  
County = 64      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check*

*0.100 AC  
+acetone*

Form 106-I8000

*80-004936  
P. 45/49 KAN*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004936  
Location = MAFB                    8164.13.00 06/09  
05/06/2010                            11:43

Test	AC	Time
01 Diagnostic	OK	11:44
02 Room Air	RFI*	11:45
03 Room Air	0.000	11:45

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Simul Ser No = 00000  
Std Sol No = 000  
County = 64                            Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*RFI Test*

Form 106-18000

80-004936  
8.44/49 1992

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004936  
Location = MAFB      8164.13.00 06/09  
05/06/2010      10:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:00
02 Std. Gas	0.076	11:00
03 Room Air	0.000	11:01
04 Std. Gas	0.077	11:01
05 Room Air	0.000	11:02
06 Std. Gas	0.077	11:02
07 Room Air	0.000	11:03

Lot No = 659358  
Cyl No = 49  
Exp Date = 09/03/2011  
County = 64      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

80-004936  
P. 4/7/99 192

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004936  
Location = MAFB      8164.13.00 06/09  
05/06/2010      11:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:04
02 Std. Gas	0.078	11:05
03 Room Air	0.000	11:05
04 Std. Gas	0.078	11:06
05 Room Air	0.000	11:06
06 Std. Gas	0.078	11:07
07 Room Air	0.000	11:07

Lot No = 659358  
Cyl No = 49  
Exp Date = 09/03/2011  
County = 64      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

80-004936  
P. 48/49 FK



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004936  
Location = MAFB      8164.13.00 06/09  
05/06/2010      11:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:08
02 Std. Gas	0.078	11:08
03 Room Air	0.000	11:09
04 Std. Gas	0.078	11:09
05 Room Air	0.000	11:10
06 Std. Gas	0.078	11:10
07 Room Air	0.000	11:11

Lot No = 659358

Cyl No = 49

Exp Date = 09/03/2011

County = 64

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

80-004936  
8.49/99 STM