



# INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59281 (5-2011)

Serial Number <b>80-004200</b>	Instrument Location <b>HETTINGER COUNTY S.O. (MOTT)</b>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <b>INSTALL DURING ANNUAL INSPECTION</b>	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000. **NA - SHARE W/FAX**
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P). ✓
  - B. ACA Test (Level 1, Function C). ✓
  - C. RFI Test (Use CMS Mode; Key Radio During First Room Air). ✓
- 7. Repair and/or Maintenance Performed (if Any): **INSTALL DURING ANNUAL INSPECTION**
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <b>CHARLES. E. EDER</b>	Date <b>06 MAY 13</b>
Reviewed By (Crime Laboratory Use Only) <b>Deb Kasher</b>	Date <b>07 May 13</b>

This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Certified By: **Charles E. Eder**      Certified Date: **29 Nov. 2016**

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004200  
Location = MOTT      8164.13.00 06/09  
05/06/2013      10:43

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?


Current Instrument Setup  
Data Entry Mode:      Enabled  
Start Test Sequence:      DABACABA  
Display Prelim Rslt?      Yes  
Display Third Digit?      Yes  
Inhib Printer(Y/N)?      No  
Display Volume?      No  
Disable On Memfull?      Yes  
# of Print Copies?      1  
Select Std (D/W/I)?      Dry  
Standard Value?      0.080  
Standard Lot #?      23411080A1  
Standard Cyl #?      60  
Standard Expiration?      10/01/2013  
Oper No?      666666

Flow Cal. Date:      12/10/2012  
Slope      637  
Intercept      -382487

IR Calibration Date:      12/10/2012

	3um	9um
0th Coef(*100):	-14942	-22738
1st Coef(*100):	256196	136921
2nd Coef(*100):	3535	1396
H2O adj(mg/l*10k):	492	515

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
Operator Signature  
CHARLES EDER

Remarks:      PRINT TEST

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004200  
Location = MOTT      8164.13.00 06/09  
05/06/2013      10:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:45
02 Std. Gas	0.080	10:46
03 Room Air	0.000	10:46
04 Std. Gas	0.080	10:46
05 Room Air	0.000	10:47
06 Std. Gas	0.080	10:47
07 Room Air	0.000	10:48

Lot No = 23411080A1  
Cyl No = 60  
Exp Date = 10/01/2013  
County = 21      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:      *ACA TEST*  
                  *0.080 AC (E.G.S.)*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004200  
Location = MOTT      8164.13.00 06/09  
05/06/2013      10:51

Test	AC	Time
01 Diagnostic	OK	10:52
02 Room Air	RFI*	10:52
03 Room Air	0.000	10:53

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = QWERTYUIOPASDFG  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 60  
Expiration Date = 10/01/2013  
County = 21      Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *RFI TEST*