

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005363

Location: DKPD

- Check When Done
- A. Is the breath tube heated? ✓
- B. Is the diagnostic check complete? ✓
- C. Is the time, date, and year correct (re-set if necessary)? ✓
- D. Print test. (Attach test record.) ✓
- E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5191 Lot No. 15160 Rep AC 0.020 ✓
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3847 Lot No. 116060 Rep AC 0.300 ✓
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5190 Lot No. FCS4 Solution 0.10AC + 0.05% acetone ✓
- H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 29614080A2 Cyl No. 10 Exp. Date 12/05/2016

Test 1 0.079 Test 4 0.080 Test 7 0.079

Test 2 0.079 Test 5 0.078 Test 8 0.078

Test 3 0.078 Test 6 0.078 Test 9 0.079

Average 0.079

J. Tank pressure, Level 3 Func D Sub G P: 326 psi Reg 320 psi

K. Remarks/Maintenance record: Annual inspection, replaced
breath hose o-ring.

06 May 16
Date
5/6/2016 ✓
Date

Deb Shonaver
Field Inspector's Signature
Charles E. [Signature]
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005363
Location = DKPD 8164.13.00 06/09
05/06/2016 08:08

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 29614080A2
Standard Cyl #? 10
Standard Expiration? 12/05/2016
Oper No? 777777

Flow Cal. Date: 08/19/2011
Slope 660
Intercept -560671

IR Calibration Date: 04/08/2015
 3um 9um

0th Coef(*100):	-12875	-23244
1st Coef(*100):	269300	140223
2nd Coef(*100):	4133	1487
H2O adj(mg/l*10k):	341	408

***** Printer Test End *****



Operator Signature
DEB SHANAVER

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005363
Location = DKPD 8164.13.00 06/09
05/06/2016 08:28

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:28
02 Std. Sol.	0.017	08:29
03 Room Air	0.000	08:30
04 Std. Sol.	0.017	08:30
05 Room Air	0.000	08:31
06 Std. Sol.	0.018	08:32
07 Room Air	0.000	08:32

08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 15160

County = 45

Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Low AC 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005363
Location = DKPD 8164.13.00 06/09
05/06/2016 08:39

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:39
02 Std. Sol.	0.292	08:40
03 Room Air	0.000	08:41
04 Std. Sol.	0.296	08:41
05 Room Air	0.000	08:42
06 Std. Sol.	0.297	08:43
07 Room Air	0.000	08:43

08 Sim Temp = 34.0°C

Simul Ser No = DR3847
Std Sol No = 16060
County = 45 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

High AC 0.300Ae

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005363
Location = DKPD 8164.13.00 06/09
05/06/2016 08:22

Test	AC	Time
01 Room Air	0.000	08:23
02 *Subject Test	INT*	08:24
03 Room Air	0.000	08:24

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 10
Expiration Date = 12/05/2016
County = 45 Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Interferent detected 0.10AC + 0.05% acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005363
Location = DKPD 8164.13.00 06/09
05/06/2016 08:25

Test	AC	Time
01 Diagnostic	OK	08:26
02 Room Air	RFI*	08:26
03 Room Air	0.000	08:27

*Invalid Test
 Inhibited - RFI


Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 10
Expiration Date = 12/05/2016
County = 45 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB SHANAVER

Remarks:




Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005363
Location = DKPD 8164.13.00 06/09
05/06/2016 08:09

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:10
02 Std. Gas	0.079	08:10
03 Room Air	0.000	08:11
04 Std. Gas	0.079	08:11
05 Room Air	0.000	08:11
06 Std. Gas	0.078	08:12
07 Room Air	0.000	08:12

Lot No = 29614080A2
Cyl No = 10
Exp Date = 12/05/2016
County = 45 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

Calibration check 0.080AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005363
Location = DKPD 8164.13.00 06/09
05/06/2016 08:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:13
02 Std. Gas	0.080	08:13
03 Room Air	0.000	08:14
04 Std. Gas	0.078	08:14
05 Room Air	0.000	08:15
06 Std. Gas	0.078	08:15
07 Room Air	0.000	08:16

Lot No = 29614080A2
Cyl No = 10
Exp Date = 12/05/2016
County = 45 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:
Calibration check 0.080AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005363
Location = DKPD 8164.13.00 06/09
05/06/2016 08:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:17
02 Std. Gas	0.079	08:17
03 Room Air	0.000	08:18
04 Std. Gas	0.078	08:18
05 Room Air	0.000	08:19
06 Std. Gas	0.079	08:19
07 Room Air	0.000	08:19

Lot No = 29614080A2
Cyl No = 10
Exp Date = 12/05/2016
County = 08 Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Calibration check 0.080 AC

Form 106-I8000