



OFFICE OF ATTORNEY GENERAL  
 Crime Laboratory Division  
 2641 East Main Avenue  
 P.O. Box 937  
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

**ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name:     Aaron R. Hummel    

Location:     GFSO    

Intoxilyzer Serial No.:     80-003057    

Lot Number:     659357    

Cylinder Number:     26    

Expiration Date:     9/3/11    

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.082%

Test 2: 0.081%

Test 3: 0.081%

(ATTACH TEST RECORD)

    Aaron R. Hummel      
 Chemical Test Operator Signature

    5/7/10 - ACA Only      
 Date Cylinder Changed

    Deb Kashner      
 Reviewed By

    11 May 2010      
 Date

Form 105-G  
 Toxicology Laboratory  
 Crime Laboratory Division

80-003057  
 P. 90-91  
 JCM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003057  
Location = GFSO      8164.13.00 06/09  
05/07/2010      17:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:09
02 Std. Gas	0.082	17:09
03 Room Air	0.000	17:10
04 Std. Gas	0.081	17:10
05 Room Air	0.000	17:11
06 Std. Gas	0.081	17:11
07 Room Air	0.000	17:12

Lot No = 659357  
Cyl No = 26  
Exp Date = 09/03/2011  
County = 18      Oper No. = 049907



Operator Signature  
AARON HUMMEL

Remarks:

ACA Test

Form 106-I8000

80-003057  
p. 91-91  
KHM