






Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004203  
Location = RUGB      8164.13.00 06/09  
05/07/2011      14:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:02
02 Std. Gas	0.082	14:02
03 Room Air	0.000	14:03
04 Std. Gas	0.081	14:03
05 Room Air	0.000	14:04
06 Std. Gas	0.081	14:04
07 Room Air	0.000	14:05

Lot No = 659358  
Cyl No = 21  
Exp Date = 09/03/2011  
County = 35      Oper No. = 069108

  
Operator Signature  
ROBERT W. KENNEDY

Remarks:

Form 106-I8000

80-004203  
P. 120 of 121  
RM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004203  
Location = RUGB                    8164.13.00 06/09  
05/07/2011                            14:05

Test	AC	Time
01 Diagnostic	OK	14:07
02 Room Air	RFI*	14:08
03 Room Air	RFI*	14:08

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, SUBJECT NA  
Sub DOB = 01/01/1955  
Sub Sex = Male                            Weight = 212  
Test = DUI                                    Cit = 888888  
Dr. Lic. = ND/TES557867  
Lot No = 659358  
Cyl No = 21  
Expiration Date = 09/03/2011  
County = 35                                    Oper No. = 069108

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
ROBERT W. KENNEDY

Remarks:

Form 106-I8000

80-004203  
p. 121 of 121  
KM