

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004188

Location: BOWMAN (BOWM)

- |  | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes?   | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?  | <u>✓</u>        |
| C. Is the breath tube heated?  | <u>✓</u>        |
| D. Is the diagnostic check complete?   | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?  | <u>✓</u>        |
| F. Print test. (Attach test record.)   | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3455</u> Lot No. <u>12210</u> Rep AC <u>0.020</u>  | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3847</u> Lot No. <u>13020</u> Rep AC <u>0.300</u>   | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR5191</u> Lot No. <u>ICST#1</u> Rep AC <u>0.100AC+0.05% ACETONE</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)  | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)  |                 |

Lot No. <u>23411080A1</u>	Cyl No. <u>69</u>	Exp. Date <u>10/01/2013</u>
Test 1 <u>0.080</u>	Test 4 <u>0.080</u>	Test 7 <u>0.080</u>
Test 2 <u>0.079</u>	Test 5 <u>0.080</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.080</u>	Test 9 <u>0.080</u>
Average <u>0.080</u>		

L. Tank pressure, Level 3 Func D Sub G P: 865 psi Reg 875 psi

M. Remarks/Maintenance record: ANNUAL INSPECTION  
REPLACED O-RING ON BREATH TUBE & SIMULATOR RETURN

06 MAY 13  
Date

Charles E. Ed...  
Field Inspector's Signature

07 May 13  
Date

Deb Kashur  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004188  
Location = BOWM                    8164.13.00 06/09  
05/06/2013                            13:57

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            23411080A1  
Standard Cyl #?            69  
Standard Expiration?        10/01/2013  
Oper No?                    666666

Flow Cal. Date:            07/05/2011  
Slope                      668  
Intercept                  -597204

IR Calibration Date:        04/05/2012  
                                  3um                    9um

-----  
0th Coef(\*100):            -15090                    -16406  
1st Coef(\*100):            263383                    126966  
2nd Coef(\*100):            6533                      2371  
H2O adj(mg/l\*10k):        801                        809

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks: PRINT TEST

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004188  
Location = BOWM      8164.13.00 06/09  
05/06/2013      14:07

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:07
02 Std. Sol.	0.019	14:08
03 Room Air	0.000	14:09
04 Std. Sol.	0.020	14:09
05 Room Air	0.000	14:10
06 Std. Sol.	0.020	14:11
07 Room Air	0.000	14:11


08 Sim Temp = 34.0°C

Simul Ser No = DR3455

Std Sol No = 12210

County = 06

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Low AC*  
*0.020 AC (sim. sol.)*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004188  
Location = BOWM      8164.13.00 06/09  
05/06/2013      14:12

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:13
02 Std. Sol.	0.303	14:13
03 Room Air	0.000	14:14
04 Std. Sol.	0.305	14:15
05 Room Air	0.000	14:15
06 Std. Sol.	0.305	14:16
07 Room Air	0.000	14:16

08 Sim Temp = 34.0°C

Simul Ser No = DR3847

Std Sol No = 13020

County = 06

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*LINEARITY TEST*

*0.300 AC (Sim. Sol.)*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004188  
Location = BOWM      8164.13.00 06/09  
05/06/2013      14:01

Test	AC	Time
01 Room Air	0.000	14:02
02 *Subject Test	INT*	14:03
03 Room Air	0.000	14:04

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = MNBVCXZLKJH1234  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 69  
Expiration Date = 10/01/2013  
County = 06      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: 0.100AC + 0.05% ACETONE

INTERFERENT CHECK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004188  
Location = BOWM                    8164.13.00 06/09  
05/06/2013                            13:59

Test	AC	Time
01 Diagnostic	OK	14:00
02 Room Air	RFI*	14:00
03 Room Air	0.000	14:01

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = QWERTYUIOPASDFG  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 69  
Expiration Date = 10/01/2013  
County = 06                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *RFI CHECK*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004188  
Location = BOWM      8164.13.00 06/09  
05/06/2013      14:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:19
02 Std. Gas	0.080	14:20
03 Room Air	0.000	14:20
04 Std. Gas	0.079	14:20
05 Room Air	0.000	14:21
06 Std. Gas	0.080	14:21
07 Room Air	0.000	14:22

Lot No = 23411080A1  
Cyl No = 69  
Exp Date = 10/01/2013  
County = 06      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC (E.G.S.)

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004188  
Location = BOWM      8164.13.00 06/09  
05/06/2013      14:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:23
02 Std. Gas	0.080	14:23
03 Room Air	0.000	14:23
04 Std. Gas	0.080	14:24
05 Room Air	0.000	14:24
06 Std. Gas	0.080	14:25
07 Room Air	0.000	14:25


Lot No = 23411080A1

Cyl No = 69

Exp Date = 10/01/2013

County = 06

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC (E.G.S.)

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004188  
Location = BOWM      8164.13.00 06/09  
05/06/2013      14:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:26
02 Std. Gas	0.080	14:26
03 Room Air	0.000	14:27
04 Std. Gas	0.080	14:27
05 Room Air	0.000	14:28
06 Std. Gas	0.080	14:28
07 Room Air	0.000	14:28

Lot No = 23411080A1

Cyl No = 69

Exp Date = 10/01/2013

County = 06

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080 AC (E.G.S.)

Form 106-I8000