



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59281 (5-2011)

Serial Number <i>80-004205-</i>	Instrument Location <i>Hill</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. RFI Test (Use CMS Mode; Key Radio During First Room Air).
- 7. Repair and/or Maintenance Performed (if Any): *other Intox sent to State Lab*
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Rae W. Weber</i>	Date <i>5/7/2015</i> ✓
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date <i>13 May 15</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = HILL 8164.13.00 06/09
05/06/2015 23:52

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 33913080A4
Standard Cyl #? 22
Standard Expiration? 01/01/2016
Oper No? 058405

Flow Cal. Date: 06/04/2009
Slope 704
Intercept -610062

IR Calibration Date: 04/17/2015
 3um 9um

0th Coef(*100): -26411 -27902
1st Coef(*100): 266995 129434
2nd Coef(*100): 2919 1480
H2O adj(mg/l*10k): 363 481

***** Printer Test End *****



Operator Signature
RAY WEBER

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = HILL 8164.13.00 06/09
05/06/2015 23:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	23:55
02 Std. Gas	0.077	23:55
03 Room Air	0.000	23:56
04 Std. Gas	0.076	23:56
05 Room Air	0.000	23:57
06 Std. Gas	0.076	23:57
07 Room Air	0.000	23:58

Lot No = 33913080A4
Cyl No = 22
Exp Date = 01/01/2016
County = 49 Oper No. = 058405



Operator Signature
RAY WEBER

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = HILL 8164.13.00 06/09
05/06/2015 23:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	23:59
02 Room Air	RFI*	23:59

*Invalid Test
Inhibited - RFI

Lot No = 33913080A4
Cyl No = 22
Exp Date = 01/01/2016
County = 49 Oper No. = 058405



Operator Signature
RAY WEBER

Remarks: