



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59281 (03-2014)

Serial Number <i>80-005948</i>	Instrument Location <i>Walsh Co SO - jail</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): *install during annual inspect*
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Deb Kashner</i>	Date <i>07 May 15 ✓</i>
Reviewed By (Crime Laboratory Use Only) <i>Charla E. Edge</i>	Date <i>11 MAY 15</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005948
Location = GRAF 8164.13.00 06/09
05/07/2015 14:21

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-==|
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-==|
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 13314080A2
Standard Cyl #? 55
Standard Expiration? 06/05/2016
Oper No? 777777

Flow Cal. Date: 06/20/2013
Slope 701
Intercept -716431

IR Calibration Date: 04/02/2015
3um 9um

0th Coef(*100): -25939 -20035
1st Coef(*100): 272687 134661
2nd Coef(*100): 3113 1576
H2O adj(mg/l*10k): 531 397

***** Printer Test End *****



Operator Signature
DEB KASHUR

Remarks:



Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005948
Location = GRAF 8164.13.00 06/09
05/07/2015 14:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:35
02 Std. Gas	0.078	14:35
03 Room Air	0.000	14:36
04 Std. Gas	0.078	14:36
05 Room Air	0.000	14:36
06 Std. Gas	0.078	14:37
07 Room Air	0.000	14:37

Lot No = 13314080A2
Cyl No = 55
Exp Date = 06/05/2016
County = 50 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005948
Location = GRAF 8164.13.00 06/09
05/07/2015 14:38

Test	AC	Time
01 Diagnostic	OK	14:38
02 Room Air	RFI*	14:38
03 Room Air	0.000	14:39

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 13314080A2
Cyl No = 55
Expiration Date = 06/05/2016
County = 50 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-I8000