

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

**ETHANOL BREATH STANDARD CYLINDER REPORT**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Steven R Fischer

Location: Rochester County Sheriff's Office, Rolla, NH

Intoxilyzer Serial No.: 80 - 004199

Lot Number: 659358

Cylinder Number: 11

Expiration Date: 9-3-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 082 %

Test 2: 0. 082 %

Test 3: 0. 081 %

(ATTACH TEST RECORD)

[Signature]  
Chemical Test Operator Signature

5-8-10  
Date

Deb Kashner  
Reviewed By

12 May 2010  
Date

80-004199  
P. 76 of 77  
KH

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004199  
Location = ROLL      8164.13.00 06/09  
05/08/2010      20:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:18
02 Std. Gas	0.082	20:19
03 Room Air	0.000	20:19
04 Std. Gas	0.082	20:20
05 Room Air	0.000	20:20
06 Std. Gas	0.081	20:21
07 Room Air	0.000	20:21

Lot No = 659358

Cyl No = 11

Exp Date = 09/03/2011

County = 40

Oper No. = 049904



Operator Signature  
STEVEN R. FISCHER

Remarks:

Form 106-I8000

80-004199  
P. 77 of 77  
YR