

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004944

Location: TOXL

- Check When Done
- A. Is the warm-up time less than 20 minutes? ✓
 - B. Is a three-pronged grounded outlet used? ✓
 - C. Is the breath tube heated? ✓
 - D. Is the diagnostic check complete? ✓
 - E. Is the time, date, and year correct (re-set if necessary)? ✓
 - F. Print test. (Attach test record.) ✓
 - G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR3455 Lot No. 12210 Rep AC 0.020 ✓
 - H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3847 Lot No. 13020 Rep AC 0.300 ✓
 - I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3377 Lot No. ICS#1 Rep AC 0.100AC + 0.05% ACETONE ✓
 - J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
 - K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 23411080A1 Cyl No. 46 Exp. Date 10/01/2013
Test 1 0.081 Test 4 0.081 Test 7 0.081
Test 2 0.081 Test 5 0.081 Test 8 0.082
Test 3 0.081 Test 6 0.081 Test 9 0.081
Average 0.081

L. Tank pressure, Level 3 Func D Sub G

P: 975 psi Reg 990 psi

M. Remarks/Maintenance record:

ANNUAL INSPECTION, REPLACED O-RING SIMULATOR RETURN
EXTERNAL BATTERY REPLACED BY DK 06 MAY 13

08 MAY 13
Date

Charles E. Eder
Field Inspector's Signature

08 May 13
Date

Deb Kashner
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
05/08/2013 09:27

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 46
Standard Expiration? 10/01/2013
Oper No? 666666

Flow Cal. Date: 07/01/2011
Slope 685
Intercept -634740

IR Calibration Date: 07/01/2011
 3um 9um

0th Coef(*100): -22591 -17234
1st Coef(*100): 265915 132691
2nd Coef(*100): 2841 1562
H2O adj(mg/l*10k): 559 425

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks: **PRINT TEST**

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
05/08/2013 09:31

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:31
02 Std. Sol.	0.019	09:32
03 Room Air	0.000	09:33
04 Std. Sol.	0.019	09:33
05 Room Air	0.000	09:34
06 Std. Sol.	0.019	09:35
07 Room Air	0.000	09:35

08 Sim Temp = 34.0°C

Simul Ser No = DR3455

Std Sol No = 12210

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC (Sim. Sol.)

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
05/08/2013 09:44

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:44
02 Std. Sol.	0.303	09:45
03 Room Air	0.000	09:46
04 Std. Sol.	0.305	09:46
05 Room Air	0.000	09:47
06 Std. Sol.	0.305	09:48
07 Room Air	0.000	09:48

08 Sim Temp = 34.0°C

Simul Ser No = DR3847

Std Sol No = 13020

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *LINER-CEE LINEARITY TEST*
0.300AC (SIM. SOL.)


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
05/08/2013 09:55

Test	AC	Time
01 Room Air	0.000	09:56
02 *Subject Test	INT*	09:57
03 Room Air	0.000	09:58

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = QWERTYUIOPASDFG
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 46
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT CHECK*
0.100AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
05/08/2013 10:00

Test	AC	Time
01 Diagnostic	OK	10:01
02 Room Air	RFI*	10:01
03 Room Air	0.000	10:02

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = MNBVCXZLKJH1234
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 46
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
05/08/2013 10:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:03
02 Std. Gas	0.081	10:03
03 Room Air	0.000	10:04
04 Std. Gas	0.081	10:04
05 Room Air	0.000	10:05
06 Std. Gas	0.081	10:05
07 Room Air	0.000	10:06

Lot No = 23411080A1
Cyl No = 46
Exp Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC (E.G.S.)


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
05/08/2013 10:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:32
02 Std. Gas	0.081	10:32
03 Room Air	0.000	10:33
04 Std. Gas	0.081	10:33
05 Room Air	0.000	10:33
06 Std. Gas	0.081	10:34
07 Room Air	0.000	10:34

Lot No = 23411080A1
Cyl No = 46
Exp Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC (E.G.S.)

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
05/08/2013 10:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:35
02 Std. Gas	0.081	10:35
03 Room Air	0.000	10:36
04 Std. Gas	0.082	10:36
05 Room Air	0.000	10:37
06 Std. Gas	0.081	10:37
07 Room Air	0.000	10:38

Lot No = 23411080A1
Cyl No = 46
Exp Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: 0.080AC (E.G.S.)
CALIBRATION CHECK