



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59281 (03-2014)

Serial Number 80-004938	Instrument Location LAWGROW
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): n/a
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Wanda L. Zeno</i>	Date 05/08/2016 ✓
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date 14 May 15

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = LANG 8164.13.00 06/09
05/08/2015 15:56

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 13314080A3
Standard Cyl #? 27
Standard Expiration? 06/05/2016
Oper No? 067518

Flow Cal. Date: 01/25/2010
Slope 712
Intercept -604165

IR Calibration Date: 04/08/2015
3um 9um

0th Coef(*100): -24553 -24676
1st Coef(*100): 265364 132923
2nd Coef(*100): 3426 1611
H2O adj(mg/l*10k): 426 442

***** Printer Test End *****



Operator Signature
DAVID J. ZEIS

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = LANG 8164.13.00 06/09
05/08/2015 15:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:58
02 Std. Gas	0.080	15:59
03 Room Air	0.000	15:59
04 Std. Gas	0.079	16:00
05 Room Air	0.000	16:00
06 Std. Gas	0.080	16:00
07 Room Air	0.000	16:01

Lot No = 13314080A3
Cyl No = 27
Exp Date = 06/05/2016
County = 10 Oper No. = 067518



Operator Signature
DAVID J. ZEIS

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = LANG 8164.13.00 06/09
05/08/2015 16:05

Test	AC	Time
01 Room Air	RFI*	16:10
02 Room Air	0.000	16:10

*Invalid Test
Inhibited - RFI

Sub Name = ZE, DA J
Sub DOB = 05/04/1951
Sub Sex = Male Weight = 194
Test = DUI Cit = 100
Dr. Lic. = ND/000000000
Lot No = 13314080A3
Cyl No = 27
Expiration Date = 06/05/2016
County = 10 Oper No. = 067518



Operator Signature
DAVID J. ZEIS

Remarks: