



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59281 (03-2014)

Serial Number 80-005943	Instrument Location Pembina County Jail (CAVA)
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i>	Date 5-8-15 ✓
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date 14 May 15

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005943
Location = CAVA 8164.13.00 06/09
05/08/2015 16:25

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 13314080A3
Standard Cyl #? 33
Standard Expiration? 06/05/2016
Oper No? 060849

Flow Cal. Date: 06/20/2013
Slope 686
Intercept -735246

IR Calibration Date: 05/05/2015
 3um 9um

0th Coef(*100): -24459 -34413
1st Coef(*100): 288841 143291
2nd Coef(*100): 2315 1134
H2O adj(mg/l*10k): 355 403

***** Printer Test End *****



Operator Signature
DANIEL C. GARVENS

Remarks: *Point Test*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005943
Location = CAVA 8164.13.00 06/09
05/08/2015 16:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:26
02 Std. Gas	0.080	16:27
03 Room Air	0.000	16:27
04 Std. Gas	0.079	16:28
05 Room Air	0.000	16:28
06 Std. Gas	0.080	16:28
07 Room Air	0.000	16:29

Lot No = 13314080A3
Cyl No = 33
Exp Date = 06/05/2016
County = 34 Oper No. = 060849



Operator Signature
DANIEL C. GARVENS

Remarks: ACA Test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005943
Location = CAVA 8164.13.00 06/09
05/08/2015 16:30

Test	AC	Time
01 Room Air	RFI*	16:31
02 Room Air	0.000	16:32

*Invalid Test
Inhibited - RFI

Sub Name = RFI, TEST NA
Sub DOB = 01/01/1980
Sub Sex = Male Weight = 100
Test = DUI Cit = 0
Dr. Lic. = ND/0
Lot No = 13314080A3
Cyl No = 33
Expiration Date = 06/05/2016
County = 34 Oper No. = 060849



Operator Signature
DANIEL C. GARVENS

Remarks: RFI Test