



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59281 (03-2014)

Serial Number 80-005449	Instrument Location GEAF
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Willie Pelt</i>	Date 8 ^{DK} 6 May 15 ✓
Reviewed By (Crime Laboratory Use Only) <i>Debi Kashner</i>	Date 13 May 15

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005949
Location = GFAF 8164.13.00 06/09
05/08/2015 14:03

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 13314080A3
Standard Cyl #? 28
Standard Expiration? 06/05/2016
Oper No? 120011

Flow Cal. Date: 06/20/2013
Slope 692
Intercept -708987

IR Calibration Date: 04/06/2015
 3um 9um

0th Coef(*100): -24726 -28868
1st Coef(*100): 263448 133361
2nd Coef(*100): 3398 1497
H2O adj(mg/l*10k): 326 367

***** Printer Test End *****



Operator Signature
WILLIAM PETTY

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005949
Location = GFAF 8164.13.00 06/09
05/08/2015 14:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:04
02 Std. Gas	0.076	14:04
03 Room Air	0.000	14:05
04 Std. Gas	0.077	14:05
05 Room Air	0.000	14:06
06 Std. Gas	0.077	14:06
07 Room Air	0.000	14:07

Lot No = 13314080A3
Cyl No = 28
Exp Date = 06/05/2016
County = 65 Oper No. = 120011



Operator Signature
WILLIAM PETTY

Remarks: A C A T e s t

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005949
Location = GFAP 8164.13.00 06/09
05/08/2015 14:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:09
02 Std. Gas	0.077	14:09
03 Room Air	RFI*	14:09
04 Room Air	RFI*	14:09

*Invalid Test
Inhibited - RFI

Lot No = 13314080A3
Cyl No = 28
Exp Date = 06/05/2016
County = 65 Oper No. = 120011



Operator Signature
WILLIAM PETTY

Remarks:

RFI check

Form 106-I8000