



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Danielle M. Roseware</u>		
Location: <u>CARR</u>		
Intoxilyzer® Serial Number: <u>80-004994</u>		
Gas Lot Number: <u>709352</u>	Gas Cylinder Number: <u>12</u>	Expiration Date: <u>12/23/11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.679</u> %		
Test 2: <u>0.079</u> %		
Test 3: <u>0.079</u> %		
(ATTACH TEST RECORD)		

Danielle M. Roseware
Chemical Test Operator Signature

05/09/10
Date

Deb Kashner
Reviewed By

10 May 2010
Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = CARR 8164.13.00 06/09
05/04/2010 08:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:21
02 Std. Gas	0.079	08:22
03 Room Air	0.000	08:22
04 Std. Gas	0.079	08:22
05 Room Air	0.000	08:23
06 Std. Gas	0.079	08:23
07 Room Air	0.000	08:24

Lot No = 709352
Cyl No = 12
Exp Date = 12/23/2011
County = 16 Oper No. = 100754



Operator Signature
DANIELLE M. ROSEWREN

Remarks:

Form 106-I8000

80-004944
P. 39/39 JH