



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Barry Vannatta</u>		
Location: <u>MobL</u>		
Intoxilyzer® Serial Number: <u>80-004947</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>27</u>	Expiration Date: <u>9-3-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>080</u> %		
Test 2: 0. <u>079</u> %		
Test 3: 0. <u>079</u> %		
(ATTACH TEST RECORD)		

Barry Vannatta  
Chemical Test Operator Signature

5-9-11  
Date

Deb Kashur  
Reviewed By

16 May 2011  
Date


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004947  
Location = MOHL      8164.13.00 06/09  
05/09/2011      17:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:00
02 Std. Gas	0.080	18:00
03 Room Air	0.000	18:00
04 Std. Gas	0.079	18:01
05 Room Air	0.000	18:01
06 Std. Gas	0.079	18:02
07 Room Air	0.000	18:02

Lot No = 659358  
Cyl No = 27  
Exp Date = 09/03/2011  
County = 38      Oper No. = 039909

  
Operator Signature  
BARRY A. VANNATTA

Remarks:

Form 106-I8000

80-004947  
P. 78 of 78  
VM