

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004199

Location: TOXL

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5144 Lot No. — Rep AC 0.021 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3847 Lot No. — Rep AC 0.294 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3452 Lot No. — Rep AC 0.10 AC + Acetone ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. <u>15410080A1</u>	Cyl No. <u>21</u>	Exp. Date <u>7-1-12</u>
Test 1 <u>0.079</u>	Test 4 <u>0.078</u>	Test 7 <u>0.079</u>
Test 2 <u>0.079</u>	Test 5 <u>0.078</u>	Test 8 <u>0.078</u>
Test 3 <u>0.078</u>	Test 6 <u>0.079</u>	Test 9 <u>0.078</u>
Average <u>0.078</u>		

L. Tank pressure, Level 3 Func D Sub G P: 253 psi Reg 250 psi

M. Remarks/Maintenance record: Annual Inspection

5-9-11
Date

[Signature]
Field Inspector's Signature

09 May 2011
Date

[Signature]
Reviewed by

80-004199
p. 98 of 106
PM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.13.00 06/09
05/09/2011 13:14

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 15410080A1
Standard Cyl #? 21
Standard Expiration? 07/01/2012
Oper No? 555555

Flow Cal. Date: 06/04/2009
Slope 701
Intercept -580043

IR Calibration Date: 06/03/2009
 3um 9um

0th Coef(*100): -44451 -27073
1st Coef(*100): 294251 132474
2nd Coef(*100): -1375 1189
H2O adj(mg/l*10k): 315 545

***** Printer Test End *****



Operator Signature
ASHLEY AYD

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.13.00 06/09
05/09/2011 13:16

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:17
02 Std. Sol.	0.018	13:18
03 Room Air	0.000	13:18
04 Std. Sol.	0.018	13:19
05 Room Air	0.000	13:19
06 Std. Sol.	0.018	13:20
07 Room Air	0.000	13:21


08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: *Low AC 0.021 AC*

Form 106-I8000

*80-004199
p. 100 of 106
km*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.13.00 06/09
05/09/2011 13:30

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:30
02 Std. Sol.	0.292	13:31
03 Room Air	0.000	13:32
04 Std. Sol.	0.295	13:33
05 Room Air	0.000	13:33
06 Std. Sol.	0.295	13:34
07 Room Air	0.000	13:34


08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: *Linearity test 0.294 AC*

Form 106-I8000

*80-004199
P. 101 of 106
KM*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.13.00 06/09
05/09/2011 13:36

Test	AC	Time
01 Room Air	0.000	13:37
02 *Subject Test	INT*	13:37
03 Room Air	0.000	13:38

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 15410080A1
Cyl No = 21
Expiration Date = 07/01/2012
County = 08 Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: *Interferent check 0.10AC + Acetone*

Form 106-I8000

80-004199
p.102 of 106
km

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.13.00 06/09
05/09/2011 13:41

Test	AC	Time
01 Diagnostic	OK	13:42
02 Room Air	RFI*	13:42
03 Room Air	0.000	13:43

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 15410080A1
Cyl No = 21
Expiration Date = 07/01/2012
County = 08 Oper No. = 555555

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ASHLEY AYD

Remarks: RFI

Form 106-I8000

80-004199
P. 103 of 106
VFM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.13.00 06/09
05/09/2011 13:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:44
02 Std. Gas	0.079	13:45
03 Room Air	0.000	13:45
04 Std. Gas	0.079	13:46
05 Room Air	0.000	13:46
06 Std. Gas	0.078	13:46
07 Room Air	0.000	13:47

Lot No = 15410080A1
Cyl No = 21
Exp Date = 07/01/2012
County = 08 Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: Calibration Check 0.080 AC

Form 106-I8000

80-004199
P.104 #106
FM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.13.00 06/09
05/09/2011 13:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:52
02 Std. Gas	0.078	13:53
03 Room Air	0.000	13:53
04 Std. Gas	0.078	13:53
05 Room Air	0.000	13:54
06 Std. Gas	0.079	13:54
07 Room Air	0.000	13:55

Lot No = 15410080A1
Cyl No = 21
Exp Date = 07/01/2012
County = 08 Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: Calibration check 0.080AC

Form 106-I8000

80-004199
P.105 of 106
KM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.13.00 06/09
05/09/2011 13:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:56
02 Std. Gas	0.079	13:56
03 Room Air	0.000	13:57
04 Std. Gas	0.078	13:57
05 Room Air	0.000	13:57
06 Std. Gas	0.078	13:58
07 Room Air	0.000	13:58

Lot No = 15410080A1
Cyl No = 21
Exp Date = 07/01/2012
County = 08 Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: Calibration check 0.080AC

Form 106-I8000

80-004199
P. 106 of 106
FM