

Certificate of Calibration

This is to certify the calibration of **Intoxilyzer**® serial number 80-005954, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to conform to the National Highway Traffic Safety Administration (NHTSA) Standard for Devices to Measure Breath Alcohol (Federal Register, Vol.58, No.179, pp 48705-48710, Sept. 17, 1993) for accuracy and precision. Reference materials are traceable through the National Institute of Standards and Technology (NIST) to the International System of Units (SI).

Date May 9, 2014 Signed [Signature]
Technician

EMI INC.

316 East 9th Street
Owensboro, KY 42303
USA

Part No. 650517 Rev.A



CMI Calibration Laboratory

Certificate of Calibration

This is to certify the breath alcohol measuring instrument, serial number 80-005954 after completion of the calibration was found to meet or exceed the acceptance criteria for accuracy and precision. This calibration was performed using reference standards and materials which are *International System of Units (SI)* traceable through the *National Institute of Standards and Technology (NIST)*.

Calibration Record Identifier CR-80-005954-334008D

Estimated Measurement Uncertainty

Breath Alcohol concentrations of:

$0 \leq 0.100 \text{ g/210L} \pm \underline{0.0020} \text{ g/210L BAC}$

$> 0.100 \text{ g/210L} \pm \underline{0.0020} \text{ g/210L BAC}$

Original

Reproduction Initials _____ Date _____

Amended Initials _____ Date _____

The expanded uncertainties estimated at the time of the calibration have been calculated using the root sum square method including both Type A and Type B components and are based on a minimum coverage factor of k=2 corresponding to an approximate confidence level of 95%.

Calibrated by: Debra Rohrer

Debra Rohrer

Date: May 12, 2014



*Breath Alcohol Measuring
Instrument Calibration
ISO/IEC 17025:2005
An ASCLD/LAB-International
Accredited Laboratory
Since February 07, 2012
Certificate # ALI-007-C*

This certificate shall not be reproduced except in full, without written approval of CMI Calibration Laboratory.