



ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Wesley Straight</i>		
Location <i>Griggs County SD</i>	Intoxilyzer® Serial Number <i>80-005940</i>	
Gas Lot Number <i>19615080A5</i>	Gas Cylinder Number <i>019</i>	Gas Expiration Date <i>9/5/17</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.080* AC
 - B. *0.080* AC
 - C. *0.077* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Wesley Straight</i>	Date <i>5/9/16</i> ✓
Reviewed By (Crime Laboratory Use Only) <i>Deb Shanaver</i>	Date <i>09 May 16</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = COOP 8164.13.00 06/09
05/09/2016 14:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:54
02 Std. Gas	0.080	14:54
03 Room Air	0.000	14:55
04 Std. Gas	0.080	14:55
05 Room Air	0.000	14:56
06 Std. Gas	0.077	14:56
07 Room Air	0.000	14:57

Lot No = 19615080A5
Cyl No = 19
Exp Date = 09/05/2017
County = 20 Oper No. = 061123



Operator Signature
WESLEY STRAIGHT

Remarks: