



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	David Feland				
Location:	Kidder S.O.				
Intoxilyzer® Serial Number:	80-004186				
Gas Lot Number:	659357	Gas Cylinder Number:	37	Expiration Date:	09/03/11
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed				
Instrument Test Results (Report to 3 Digits; Example, 0.081%)					
Test 1: <u>0.080</u> %					
Test 2: <u>0.079</u> %					
Test 3: <u>0.080</u> %					
(ATTACH TEST RECORD)					

David Feland
Chemical Test Operator Signature

10 May 10
Date

Deb Kashner
Reviewed By

17 May 2010
Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
05/10/2010 10:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:16
02 Std. Gas	0.080	10:16
03 Room Air	0.000	10:17
04 Std. Gas	0.079	10:17
05 Room Air	0.000	10:17
06 Std. Gas	0.080	10:18
07 Room Air	0.000	10:18

Lot No = 659357
Cyl No = 37
Exp Date = 09/03/2011
County = 22 Oper No. = 119404



Operator Signature
DAVID R. FELAND

Remarks:

Form 106-I8000

80-004186
P. 73/73 KAL