

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004186

Location: Kidder Co 50

- | | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5132</u> Lot No. <u>—</u> Rep AC <u>0.021 AC</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3375</u> Lot No. <u>—</u> Rep AC <u>0.292 AC</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR5186</u> Lot No. <u>—</u> Rep AC <u>0.10 AC + Acetone 0.05%</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. <u>659357</u>	Cyl No. <u>37</u>	Exp. Date <u>09/03/11</u>
Test 1 <u>0.077</u>	Test 4 <u>0.079</u>	Test 7 <u>0.080 0.079</u>
Test 2 <u>0.080</u>	Test 5 <u>0.079</u>	Test 8 <u>0.080</u>
Test 3 <u>0.079</u>	Test 6 <u>0.079</u>	Test 9 <u>0.080</u>
Average <u>0.079 AC</u>		

L. Tank pressure, Level 3 Func D Sub G P: 901 psi Reg 900 psi

M. Remarks/Maintenance record: Annual Inspection

10 May 2010
Date

Mary Plasm
Field Inspector's Signature

DK¹⁵ 17 May 2010
Date

Deb Kasher
Reviewed by

80-004186
P. 74/82 KR

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
05/10/2010 10:05

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 659357
Standard Cyl #? 37
Standard Expiration? 09/03/2011
Oper No? 107501

Flow Cal. Date: 06/03/2009
Slope 703
Intercept -668211

IR Calibration Date: 06/03/2009
 3um 9um

0th Coef(*100): -12098 -31908
1st Coef(*100): 257966 136312
2nd Coef(*100): 2942 1009
H2O adj(mg/l*10k): 434 609

***** Printer Test End *****



Operator Signature
MARGY PEARSON

Remarks:

Print

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
05/10/2010 09:41

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:42
02 Std. Sol.	0.016	09:43
03 Room Air	0.000	09:44
04 Std. Sol.	0.017	09:44
05 Room Air	0.000	09:45
06 Std. Sol.	0.017	09:45
07 Room Air	0.000	09:46


08 Sim Temp = 34.0°C

Simul Ser No = 000000

Std Sol No = 000000

County = 22

Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

Low AC 0.021 AC

Form 106-I8000

80-004186
P. 76/82 KA

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
05/10/2010 09:53

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:54
02 Std. Sol.	0.302	09:55
03 Room Air	0.000	09:55
04 Std. Sol.	0.301	09:56
05 Room Air	0.000	09:56
06 Std. Sol.	0.300	09:57
07 Room Air	0.000	09:58

08 Sim Temp = 34.0°C

Simul Ser No = 000000
Std Sol No = 000000
County = 22 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

Linearity 0.292 AC

Form 106-18000

80-004186
P. 77/82 PK


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
05/10/2010 09:38

Test	AC	Time
01 Room Air	0.000	09:39
02 *Subject Test	INT*	09:40
03 Room Air	0.000	09:40

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 659358
Cyl No = 43
Expiration Date = 09/03/2011
County = 22 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

Interferent Alc + Acetone

Form 106-18000

80-004186
P. 78/82 FK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
05/10/2010 09:35

Test	AC	Time
01 Diagnostic	OK	09:36
02 Room Air	RFI*	09:36
03 Room Air	0.000	09:37

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 659358
Cyl No = 43
Expiration Date = 09/03/2011
County = 22 Oper No. = 107501

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
MARGY PEARSON

Remarks:

RFI

Form 106-I8000

80-004186
P. 79/82 KA

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
05/10/2010 10:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:02
02 Std. Gas	0.077	10:02
03 Room Air	0.000	10:03
04 Std. Gas	0.080	10:03
05 Room Air	0.000	10:04
06 Std. Gas	0.079	10:04
07 Room Air	0.000	10:05

Lot No = 659357
Cyl No = 37
Exp Date = 09/03/2011
County = 22 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

Cal Check ACA

Form 106-I8000

80-004186
P. 80/82 FR


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
05/10/2010 10:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:06
02 Std. Gas	0.079	10:07
03 Room Air	0.000	10:07
04 Std. Gas	0.079	10:08
05 Room Air	0.000	10:08
06 Std. Gas	0.079	10:09
07 Room Air	0.000	10:09

Lot No = 659357
Cyl No = 37
Exp Date = 09/03/2011
County = 08 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

Cal Check ACA

Form 106-18000

80-004186
P. 5/12 KM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
05/10/2010 10:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:10
02 Std. Gas	0.079	10:11
03 Room Air	0.000	10:11
04 Std. Gas	0.080	10:12
05 Room Air	0.000	10:12
06 Std. Gas	0.080	10:13
07 Room Air	0.000	10:13

Lot No = 659357
Cyl No = 37
Exp Date = 09/03/2011
County = 22 Oper No. = 107501

Margy Pearson

Operator Signature
MARGY PEARSON

Remarks:

Cal Check ACA

Form 106-I8000

80-004186
P. 5/8/12 PM