



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	Margy Pearson		107501
Location:	UND PD		
Intoxilyzer® Serial Number:	80-004934		
Gas Lot Number:	659358	Gas Cylinder Number:	31
		Expiration Date:	9/03/2011
<input checked="" type="checkbox"/> New Cylinder Installed	<input type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)			
Test 1: 0.079%			
Test 2: 0.079%			
Test 3: 0.079%			
(ATTACH TEST RECORD)			

Margy Pearson  
 Chemical Test Operator Signature

10 May 2010  
 Date

Deb Kashner  
 Reviewed By

DK<sup>15</sup> 17 May 2010  
 Date

80-004934  
 P. 4/2/13 YAM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004934  
Location = UNDP      8164.13.00 06/09  
05/10/2010      19:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:08
02 Std. Gas	0.079	19:08
03 Room Air	0.000	19:09
04 Std. Gas	0.079	19:09
05 Room Air	0.000	19:10
06 Std. Gas	0.079	19:10
07 Room Air	0.000	19:10

Lot No = 659358  
Cyl No = 31  
Exp Date = 09/03/2011  
County = 18      Oper No. = 107501

*Margy Pearson*

Operator Signature  
MARGY PEARSON

Remarks:

*Cyl install - ACA*

Form 106-I8000

80-004934  
P. 43/43 VAA