



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Dustin B. Newman</u>		
Location: <u>Walford City, ND</u>		
Intoxilyzer® Serial Number: <u>80-004938</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>44</u>	Expiration Date: <u>09/03/11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>079</u> %		
Test 2: 0. <u>079</u> %		
Test 3: 0. <u>080</u> %		
(ATTACH TEST RECORD)		

Dustin B. Newman
Chemical Test Operator Signature

05/10/11
Date

Deb Kashner
Reviewed By

16 May 2011
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = WATF 8164.13.00 06/09
05/10/2011 00:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	00:00
02 Std. Gas	0.079	00:01
03 Room Air	0.000	00:01
04 Std. Gas	0.079	00:01
05 Room Air	0.000	00:02
06 Std. Gas	0.080	00:02
07 Room Air	0.000	00:03

Lot No = 659357
Cyl No = 44
Exp Date = 09/03/2011
County = 27 Oper No. = 030924



Operator Signature
DUSTIN B. NEWMAN

Remarks: *Good ACA Test*

Form 106-I8000

*80-004938
p. 87 of 87
RN*