



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Hinton, Rick W</u>		
Location: <u>Cass County Jail</u>		
Intoxilyzer® Serial Number: <u>80-004946</u>		
Gas Lot Number: <u>1540080A</u>	Gas Cylinder Number: <u>10</u>	Expiration Date: <u>7-1-12</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: 0. <u>074</u> % Test 2: 0. <u>074</u> % Test 3: 0. <u>074</u> %		
(ATTACH TEST RECORD)		

Rick W Hinton
 Chemical Test Operator Signature

5-10-11
 Date

Deb Kashner
 Reviewed By

16 May 2011
 Date

As this form indicates, the results indicate a reading not acceptable to ND law. The operation of this Intoxiler has been stopped until such time the readings fall within the required limits.

80-004946
 P. 80 of 81
 FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = CASS 8164.13.00 06/09
05/09/2011 22:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	22:45
02 Std. Gas	0.074	22:46
03 Room Air	0.000	22:46
04 Std. Gas	0.074	22:47
05 Room Air	0.000	22:47
06 Std. Gas	0.074	22:47
07 Room Air	0.000	22:48

Lot No = 15410080A1
Cyl No = 10
Exp Date = 07/01/2012
County = 09 Oper No. = 029920



Operator Signature
RICK W. HINTON

Remarks:

Form 106-I8000

80-004946
P. 8/2/81
AM