



ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION

Toxicology Section/Breath Alcohol Program

SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Deb Shanaver</i>		
Location <i>G^{DS} Walsh Co SO</i>	Intoxilyzer® Serial Number <i>80-005948</i>	
Gas Lot Number <i>01316080A2</i>	Gas Cylinder Number <i>27</i>	Gas Expiration Date <i>02/05/2018</i>

Check When Done:

1. Scan/Enter Gas Cylinder Information (Level 1, Function S).

2. Perform an ACA Test (Level 1, Function C).

Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).

A. 0. 078 AC

B. 0. 077 AC

C. 0. 078 AC

3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.

4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.

5. Send the Following to the Crime Laboratory:

A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).

B. ACA Test Record.

Chemical Test Operator Signature <i>Deb Shanaver</i>	Date <i>10 May 16 ✓</i>
Reviewed By (Crime Laboratory Use Only) <i>Charles E. El...</i>	Date <i>12 MAY 2016</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005948
Location = GRAF 8164.13.00 06/09
05/10/2016 12:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:49
02 Std. Gas	0.078	12:49
03 Room Air	0.000	12:50
04 Std. Gas	0.077	12:50
05 Room Air	0.000	12:51
06 Std. Gas	0.078	12:51
07 Room Air	0.000	12:52

Lot No = 01316080A2
Cyl No = 27
Exp Date = 02/05/2018
County = 50 Oper No. = 777777


Operator Signature
DEB SHANAVER

Remarks: *ACA*
Calibration check 0.080AC