

INTOXILYZER 8000 INSPECTION

Serial No.: 80-002669

Location: Pembina Co 50

- |   | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes?  | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?   | <u>✓</u>        |
| C. Is the breath tube heated?   | <u>✓</u>        |
| D. Is the diagnostic check complete?  | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?   | <u>✓</u>        |
| F. Print test. (Attach test record.)  | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5132</u> Lot No. <u>—</u> Rep AC <u>0.021 AC</u>  | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3375</u> Lot No. <u>—</u> Rep AC <u>0.292 AC</u>   | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR5186</u> Lot No. <u>—</u> Rep AC <u>0.10 AC + 0.05% Acetone</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)   | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)   |                 |

Lot No. <u>659357</u>	Cyl No. <u>9</u>	Exp. Date <u>09/03/2011</u>
Test 1 <u>0.077</u>	Test 4 <u>0.077</u>	Test 7 <u>0.077</u>
Test 2 <u>0.076</u>	Test 5 <u>0.077</u>	Test 8 <u>0.077</u>
Test 3 <u>0.077</u>	Test 6 <u>0.077</u>	Test 9 <u>0.077</u>
Average <u>0.077</u>		

L. Tank pressure, Level 3 Func D Sub G P: 785 psi Reg 800 psi

M. Remarks/Maintenance record: Annual Inspection

11 May 2010  
Date  
15  
17 May 2010  
Date

Mary Pearson  
Field Inspector's Signature  
Deb Kashner  
Reviewed by



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002669  
Location = CAVA      8164.13.00 06/09  
05/11/2010      12:50

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:51
02 Std. Sol.	0.017	12:52
03 Room Air	0.000	12:52
04 Std. Sol.	0.018	12:53
05 Room Air	0.000	12:54
06 Std. Sol.	0.019	12:54
07 Room Air	0.000	12:55

08 Sim Temp = 34.0°C

Simul Ser No = 00000000

Std Sol No = 00000000

County = 18

Oper No. = 107501

*Margy Pearson*

Operator Signature  
MARGY PEARSON

Remarks:

*Low AC 0.021 AC*

Form 106-I8000

*80-002669  
p. 85 of 91*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002669  
Location = CAVA      8164.13.00 06/09  
05/11/2010      12:56

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:56
02 Std. Sol.	0.292	12:57
03 Room Air	0.000	12:58
04 Std. Sol.	0.294	12:58
05 Room Air	0.000	12:59
06 Std. Sol.	0.294	13:00
07 Room Air	0.000	13:00


08 Sim Temp = 34.0°C

Simul Ser No = 00000000

Std Sol No = 00000000

County = 18

Oper No. = 107501



Operator Signature  
MARGY PEARSON

Remarks:

Linearity 0.292AC

Form 106-I8000

80-002669  
P. 8-1-91  
RM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002669  
Location = CAVA      8164.13.00 06/09  
05/11/2010      12:41

Test	AC	Time
01 Room Air	0.000	12:42
02 *Subject Test	INT*	12:43
03 Room Air	0.000	12:43

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS821456  
Simul Ser No = 00000000  
Std Sol No = 00000000  
County = 34      Oper No. = 107501

  
\_\_\_\_\_  
Operator Signature  
MARGY PEARSON

Remarks:

*Interfer 0.10 AC + 0.05% Acetone*

Form 106-I8000

80-002669  
P. 87491  
JRM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002669  
Location = CAVA      8164.13.00 06/09  
05/11/2010      12:46

Test	AC	Time
01 Diagnostic	OK	12:47
02 Room Air	RFI*	12:47
03 Room Air	0.000	12:48

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS821456  
Simul Ser No = 00000000  
Std Sol No = 00000000  
County = 18      Oper No. = 107501

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
MARGY PEARSON

Remarks:

RFI

Form 106-I8000

80-002669  
P 88 of 91  
KM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002669  
Location = CAVA      8164.13.00 06/09  
05/11/2010      13:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:08
02 Std. Gas	0.077	13:08
03 Room Air	0.000	13:09
04 Std. Gas	0.076	13:09
05 Room Air	0.000	13:10
06 Std. Gas	0.077	13:10
07 Room Air	0.000	13:11


Lot No = 659357

Cyl No = 9

Exp Date = 09/03/2011

County = 18

Oper No. = 107501



Operator Signature  
MARGY PEARSON

Remarks:

*CalCheck ACA*

Form 106-I8000

*80-002669  
P. 89/9/09*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002669  
Location = CAVA      8164.13.00 06/09  
05/11/2010      13:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:12
02 Std. Gas	0.077	13:12
03 Room Air	0.000	13:13
04 Std. Gas	0.077	13:13
05 Room Air	0.000	13:14
06 Std. Gas	0.077	13:14
07 Room Air	0.000	13:15

Lot No = 659357  
Cyl No = 9  
Exp Date = 09/03/2011  
County = 18      Oper No. = 107501

  
\_\_\_\_\_  
Operator Signature  
MARGY PEARSON

Remarks:

*Cal Check 0.080 AC*

Form 106-I8000

80-002669  
P. 90 of 91  
FM



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002669  
Location = CAVA      8164.13.00 06/09  
05/11/2010      13:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:16
02 Std. Gas	0.077	13:16
03 Room Air	0.000	13:17
04 Std. Gas	0.077	13:17
05 Room Air	0.000	13:18
06 Std. Gas	0.077	13:18
07 Room Air	0.000	13:19

Lot No = 659357

Cyl No = 9

Exp Date = 09/03/2011

County = 18

Oper No. = 107501

  
\_\_\_\_\_  
Operator Signature  
MARGY PEARSON

Remarks:

*Cal Check ACA 0.080 AC*

Form 106-I8000

80-002669  
P. 9/1/91  
SM