

INTOXILYZER 8000 INSPECTION

Serial No.: 80-003072

Location: Walsh Co SU

- |   | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes?  | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?   | <u>✓</u>        |
| C. Is the breath tube heated?   | <u>✓</u>        |
| D. Is the diagnostic check complete?  | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?   | <u>✓</u>        |
| F. Print test. (Attach test record.)  | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5186</u> Lot No. <u>-</u> Rep AC <u>0.021</u>   | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3375</u> Lot No. <u>-</u> Rep AC <u>0.292 AC</u>   | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode. (Attach test record.)<br>Sim SN <u>DR 5186</u> Lot No. <u>-</u> Rep AC <u>0.10 AC + 0.05% Acetone</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air. (Attach test record.)  | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode. (Attach test record.)  |                 |

Lot No. <u>659358</u>	Cyl No. <u>16</u>	Exp. Date <u>09/03/2011</u>
Test 1 <u>0.080</u>	Test 4 <u>0.081</u>	Test 7 <u>0.082</u>
Test 2 <u>0.082</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.082</u>	Test 6 <u>0.082</u>	Test 9 <u>0.081</u>
Average <u>0.081 AC</u>		

L. Tank pressure, Level 3 Func D Sub G MAP 703  
P: 250 psi Reg 750 psi

M. Remarks/Maintenance record: Annual Inspection

11 May 2010  
Date

Mary Pearson  
Field Inspector's Signature

pk 15 May 2010  
Date

Deb Kasher  
Reviewed by

80-003072  
8/88 KAN

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer  
North Dakota Model 8000 SN 80-003072  
Location = GRAF 8164.13.00 06/09  
05/11/2010 10:51

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On Memfull? Yes  
# of Print Copies? 1  
Select Std (D/W/I)? Dry  
Standard Value? 0.080  
Standard Lot #? 659358  
Standard Cyl #? 16  
Standard Expiration? 09/03/2011  
Oper No? 107501


Flow Cal. Date: 06/05/2008  
Slope 732  
Intercept -720104

IR Calibration Date: 06/05/2008

	3um	9um
--	-----	-----

0th Coef(*100):	-20044	-17692
1st Coef(*100):	264480	139417
2nd Coef(*100):	1329	776
H2O adj(mg/l*10k):	540	318

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
MARGY PEARSON

Remarks:

Print

Form 106-I8000

80-003072  
8/88 KAL

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003072  
Location = GRAF      8164.13.00 06/09  
05/11/2010      11:11

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:12
02 Std. Sol.	0.018	11:13
03 Room Air	0.000	11:13
04 Std. Sol.	0.018	11:14
05 Room Air	0.000	11:15
06 Std. Sol.	0.018	11:15
07 Room Air	0.000	11:16

08 Sim Temp = 34.0°C

Simul Ser No = 00000000

Std Sol No = 000000

County = 50

Oper No. = 107501

*Margy Pearson*

Operator Signature  
MARGY PEARSON

Remarks:

ACA LOW AC 0.021 AC

Form 106-I8000

80-003072  
8/88 FM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003072  
Location = GRAF      8164.13.00 06/09  
05/11/2010      11:06

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:07
02 Std. Sol.	0.290	11:08
03 Room Air	0.000	11:08
04 Std. Sol.	0.291	11:09
05 Room Air	0.000	11:09
06 Std. Sol.	0.293	11:10
07 Room Air	0.000	11:11


08 Sim Temp = 34.0°C

Simul Ser No = 00000000

Std Sol No = 000000

County = 50

Oper No. = 107501



Operator Signature  
MARGY PEARSON

Remarks:

Linearity = 0.292 AC

Form 106-18000

80-003072  
8/3/88 FAL

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003072  
Location = GRAF      8164.13.00 06/09  
05/11/2010      11:00

Test	AC	Time
01 Room Air	0.000	11:01
02 *Subject Test	INT*	11:02
03 Room Air	0.000	11:02

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS821456  
Simul Ser No = 00000000  
Std Sol No = 000000  
County = 50      Oper No. = 107501

  
\_\_\_\_\_  
Operator Signature  
MARGY PEARSON

Remarks:

*Interferent 0.10 AC + 0.05% Acetone*

Form 106-18000

80-003072  
8/1/88 FR

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003072  
Location = GRAF      8164.13.00 06/09  
05/11/2010      10:52

Test	AC	Time
01 Diagnostic	OK	10:53
02 Room Air	RFI*	10:53
03 Room Air	0.000	10:54

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS821456  
Lot No = 659358  
Cyl No = 16  
Expiration Date = 09/03/2011  
County = 50      Oper No. = 107501

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

  
\_\_\_\_\_  
Operator Signature  
MARGY PEARSON

Remarks:

RFI

Form 106-I8000

80-003072  
8/8/88 KM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003072  
Location = GRAF      8164.13.00 06/09  
05/11/2010      11:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:18
02 Std. Gas	0.080	11:18
03 Room Air	0.000	11:19
04 Std. Gas	0.082	11:19
05 Room Air	0.000	11:20
06 Std. Gas	0.082	11:20
07 Room Air	0.000	11:21

Lot No = 659358  
Cyl No = 16  
Exp Date = 09/03/2011  
County = 50      Oper No. = 107501

  
\_\_\_\_\_  
Operator Signature  
MARGY PEARSON

Remarks:

*Cal Check 0.080 AC*

Form 106-I8000

80-003072  
8/6/88 KAL


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003072  
Location = GRAF      8164.13.00 06/09  
05/11/2010      11:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:21
02 Std. Gas	0.081	11:22
03 Room Air	0.000	11:22
04 Std. Gas	0.081	11:23
05 Room Air	0.000	11:23
06 Std. Gas	0.082	11:23
07 Room Air	0.000	11:24

Lot No = 659358  
Cyl No = 16  
Exp Date = 09/03/2011  
County = 50      Oper No. = 107501

  
\_\_\_\_\_  
Operator Signature  
MARGY PEARSON

Remarks:

*Cal Check 0.080 AC*

Form 106-I8000

80-003072  
8/88 *JKL*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003072  
Location = GRAF      8164.13.00 06/09  
05/11/2010      11:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:25
02 Std. Gas	0.082	11:25
03 Room Air	0.000	11:26
04 Std. Gas	0.081	11:26
05 Room Air	0.000	11:27
06 Std. Gas	0.081	11:27
07 Room Air	0.000	11:28

Lot No = 659358  
Cyl No = 16  
Exp Date = 09/03/2011  
County = 50      Oper No. = 107501

*Margy Pearson*

Operator Signature  
MARGY PEARSON

Remarks:

*Cal Check 0.080 AC*

Form 106-I8000

80-003072  
8/18/08 *KA*