



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Lt. Tracy L. Meidinger #258</u>		
Location: <u>UND-PO</u>		
Intoxilyzer® Serial Number: <u>80-004934</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>31</u>	Expiration Date: <u>09-30-2011</u> ^{03 DK}
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>080</u> %		
Test 2: 0. <u>079</u> %		
Test 3: 0. <u>080</u> %		
(ATTACH TEST RECORD)		

Lt. Tracy L. Meidinger
Chemical Test Operator Signature

05-11-2011
Date

Deb Kashner
Reviewed By

16 May 2011
Date

