

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-003065

Location: GF Corr.

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5143 Lot No. 201501B Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3455 Lot No. 14270 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5132 Lot No. 1CS#2 Solution 0.10AC + 0.05% acetone ✓
 - H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. <u>14214080A6</u>	Cyl No. <u>52</u>	Exp. Date <u>07/05/2016</u>
Test 1 <u>0.082</u>	Test 4 <u>0.080</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.080</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>
Average <u>0.081</u>		

J. Tank pressure, Level 3 Func D Sub G 1 P: 1017 psi Reg 1000 psi

K. Remarks/Maintenance record: annual inspection, cleaned breath hose screen

08 May 15
Date

11 MAY 15 ✓
Date

Deb Kashur
Field Inspector's Signature

Charles Elden
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = GFSO 8164.13.00 06/09
05/08/2015 10:15

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 14214080A6
Standard Cyl #? 52
Standard Expiration? 07/05/2016
Oper No? 777777

Flow Cal. Date: 04/08/2008
Slope 694
Intercept -539075

IR Calibration Date: 06/05/2008
 3um 9um

0th Coef(*100): -9383 -9235
1st Coef(*100): 288128 138520
2nd Coef(*100): 957 781
H2O adj(mg/l*10k): 391 239

***** Printer Test End *****



Operator Signature
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = GFSO 8164.13.00 06/09
05/08/2015 10:27

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:27
02 Std. Sol.	0.020	10:28
03 Room Air	0.000	10:29
04 Std. Sol.	0.020	10:29
05 Room Air	0.000	10:30
06 Std. Sol.	0.020	10:31
07 Room Air	0.000	10:31

08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 201501B

County = 18

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

low AC 0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = GFSO 8164.13.00 06/09
05/08/2015 10:32

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:33
02 Std. Sol.	0.293	10:34
03 Room Air	0.000	10:34
04 Std. Sol.	0.293	10:35
05 Room Air	0.000	10:36
06 Std. Sol.	0.293	10:36
07 Room Air	0.000	10:37

08 Sim Temp = 34.0°C

Simul Ser No = DR3455
Std Sol No = 14270
County = 18 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:
High AC ^{DK} ~~0.0~~ 0.300 1e

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = GFSO 8164.13.00 06/09
05/08/2015 10:38

Test	AC	Time
01 Room Air	0.000	10:39
02 *Subject Test	INT*	10:39
03 Room Air	0.000	10:40

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 14214080A6
Cyl No = 52
Expiration Date = 07/05/2016
County = 18 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.10AC + 0.05% acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = GFSO 8164.13.00 06/09
05/08/2015 10:41

Test	AC	Time
01 Diagnostic	OK	10:42
02 Room Air	RFI*	10:42
03 Room Air	0.000	10:42

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 14214080A6
Cyl No = 52
Expiration Date = 07/05/2016
County = 18 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = GFSO 8164.13.00 06/09
05/08/2015 10:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:18
02 Std. Gas	0.082	10:19
03 Room Air	0.000	10:19
04 Std. Gas	0.081	10:19
05 Room Air	0.000	10:20
06 Std. Gas	0.080	10:20
07 Room Air	0.000	10:21

Lot No = 14214080A6
Cyl No = 52
Exp Date = 07/05/2016
County = 18 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = GFSO 8164.13.00 06/09
05/08/2015 10:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:23
02 Std. Gas	0.080	10:24
03 Room Air	0.000	10:24
04 Std. Gas	0.081	10:25
05 Room Air	0.000	10:25
06 Std. Gas	0.081	10:26
07 Room Air	0.000	10:26

Lot No = 14214080A6
Cyl No = 52
Exp Date = 07/05/2016
County = 18 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = GFSO 8164.13.00 06/09
05/08/2015 10:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:43
02 Std. Gas	0.081	10:44
03 Room Air	0.000	10:44
04 Std. Gas	0.081	10:45
05 Room Air	0.000	10:45
06 Std. Gas	0.081	10:45
07 Room Air	0.000	10:46

Lot No = 14214080A6
Cyl No = 52
Exp Date = 07/05/2016
County = 18 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000