

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004198

Location: GFPD

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5143 Lot No. 201501B Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3455 Lot No. 14276 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5132 Lot No. 1CS#2 Solution 0.10 AC + 0.05% acetone ✓
 - H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 13314080A2 Cyl No. 56 Exp. Date 06/05/2016

Test 1 <u>0.083</u>	Test 4 <u>0.083</u>	Test 7 <u>0.083</u>
Test 2 <u>0.083</u>	Test 5 <u>0.083</u>	Test 8 <u>0.083</u>
Test 3 <u>0.083</u>	Test 6 <u>0.083</u>	Test 9 <u>0.083</u>

Average 0.083

J. Tank pressure, Level 3 Func D Sub G P: 984 psi Reg 2/1000 psi

K. Remarks/Maintenance record: Annual inspection, calibrated gas regulator

08 May 2015
Date

11 MAY 15 ✓
Date

Deb Kashner
Field Inspector's Signature

Charla E. Edwards
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
05/08/2015 09:24

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 13314080A2
Standard Cyl #? 56
Standard Expiration? 06/05/2016
Oper No? 777777

Flow Cal. Date: 01/09/2012
Slope 637
Intercept -467208

IR Calibration Date: 01/09/2012
 3um 9um

0th Coef(*100): -31777 -32014
1st Coef(*100): 271021 132277
2nd Coef(*100): 1454 964
H2O adj(mg/l*10k): 558 576

***** Printer Test End *****



Operator Signature
DEB KASHUR

Remarks:



Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
05/08/2015 09:38

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:39
02 Std. Sol.	0.018	09:40
03 Room Air	0.000	09:40
04 Std. Sol.	0.019	09:41
05 Room Air	0.000	09:41
06 Std. Sol.	0.018	09:42
07 Room Air	0.000	09:42

08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 210501B

County = 18

Oper No. = 777777

Deb Kashur

Operator Signature

DEB KASHUR

Remarks:

Low AC 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
05/08/2015 09:44

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:45
02 Std. Sol.	0.298	09:45
03 Room Air	0.000	09:46
04 Std. Sol.	0.299	09:47
05 Room Air	0.000	09:47
06 Std. Sol.	0.298	09:48
07 Room Air	0.000	09:48

08 Sim Temp = 34.0°C

Simul Ser No = DR3455
Std Sol No = 14270
County = 18 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

High AC 0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
05/08/2015 09:36

Test	AC	Time
01 Room Air	0.000	09:37
02 *Subject Test	INT*	09:37
03 Room Air	0.000	09:38

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 13314080A2
Cyl No = 56
Expiration Date = 06/05/2016
County = 18 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.10AC + 0.05% acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
05/08/2015 09:49

Test	AC	Time
01 Diagnostic	OK	09:50
02 Room Air	RFI*	09:50
03 Room Air	0.000	09:51

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 13314080A2
Cyl No = 56
Expiration Date = 06/05/2016
County = 18 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
05/08/2015 09:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:24
02 Std. Gas	0.083	09:25
03 Room Air	0.000	09:25
04 Std. Gas	0.083	09:26
05 Room Air	0.000	09:26
06 Std. Gas	0.083	09:27
07 Room Air	0.000	09:27

Lot No = 13314080A2
Cyl No = 56
Exp Date = 06/05/2016
County = 18 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks: Calibration check
~~Low AC~~ 0.080 AC
DK
Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
05/08/2015 09:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:28
02 Std. Gas	0.083	09:28
03 Room Air	0.000	09:29
04 Std. Gas	0.083	09:29
05 Room Air	0.000	09:29
06 Std. Gas	0.083	09:30
07 Room Air	0.000	09:30

Lot No = 13314080A2
Cyl No = 56
Exp Date = 06/05/2016
County = 18 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = GFPD 8164.13.00 06/09
05/08/2015 09:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:33
02 Std. Gas	0.083	09:33
03 Room Air	0.000	09:34
04 Std. Gas	0.083	09:34
05 Room Air	0.000	09:35
06 Std. Gas	0.083	09:35
07 Room Air	0.000	09:35

Lot No = 13314080A2

Cyl No = 56

Exp Date = 06/05/2016

County = 18

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC

Form 106-I8000