



ETHANOL GAS STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Deb Shanaver</i>		
Location <i>Minot Air Force Base</i>	Intoxilyzer® Serial Number <i>80-004943</i>	
Gas Lot Number <i>01316080A2</i>	Gas Cylinder Number <i>30</i>	Gas Expiration Date <i>02/05/2018</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.080* AC
 B. *0.079* AC
 C. *0.079* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Deb Shanaver</i> ✓	Date <i>11 May 16</i> ✓
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <i>12 MAY 16</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = MNAF 8164.13.00 06/09
05/11/2016 13:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:42
02 Std. Gas	0.080	13:43
03 Room Air	0.000	13:43
04 Std. Gas	0.079	13:43
05 Room Air	0.000	13:44
06 Std. Gas	0.079	13:44
07 Room Air	0.000	13:45

Lot No = 01316080A2
Cyl No = 30
Exp Date = 02/05/2018
County = 64 Oper No. = 777777

Deb Shanaver
Operator Signature
DEB SHANAVER

Remarks: ACA
Calibration check 0.080AC