



ETHANOL GAS STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Deb Shanaver</i>		
Location <i>BIA Belcourt</i>	Intoxilyzer® Serial Number <i>80-004946</i>	
Gas Lot Number <i>19615080 AS</i>	Gas Cylinder Number <i>20</i>	Gas Expiration Date <i>09/05/2017</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.018* AC
 B. *0.019* AC
 C. *0.019* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Deb Shanaver</i>	Date <i>11 May 16</i> ✓
Reviewed By (Crime Laboratory Use Only) <i>Charles E. ...</i>	Date <i>12 MAY 16</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = BELC 8164.13.00 06/09
05/11/2016 09:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:54
02 Std. Gas	0.078	09:54
03 Room Air	0.000	09:54
04 Std. Gas	0.079	09:55
05 Room Air	0.000	09:55
06 Std. Gas	0.079	09:56
07 Room Air	0.000	09:56

Lot No = 19615080A5
Cyl No = 20
Exp Date = 09/05/2017
County = 63 Oper No. = 777777

Deb Shanaver
Operator Signature
DEB SHANAVER

Remarks:

ACA 0.080AC