

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004193

Location: Lake Region Corr Cen

- |  | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes?   | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?  | <u>✓</u>        |
| C. Is the breath tube heated?  | <u>✓</u>        |
| D. Is the diagnostic check complete?   | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?  | <u>✓</u>        |
| F. Print test. (Attach test record.)   | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5132</u> Lot No. <u>    </u> Rep AC <u>0.021 AC</u>  | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3375</u> Lot No. <u>    </u> Rep AC <u>0.292 AC</u>   | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR5186</u> Lot No. <u>    </u> Rep AC <u>0.10 AC + 0.05% Acetone</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)  | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)  |                 |

Lot No. <u>659358</u>	Cyl No. <u>20</u>	Exp. Date <u>09/03/2011</u>
Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.081</u>	Test 6 <u>0.081</u>	Test 9 <u>0.082</u>
Average <u>0.081</u>		

L. Tank pressure, Level 3 Func D Sub G P: 479 psi Reg 500 psi

M. Remarks/Maintenance record: Annual Inspection

12 May 2010  
Date

Margy Pearson  
Field Inspector's Signature

15  
OK 17 May 2010  
Date

Deb Kashner  
Reviewed by

80-004193  
P. 72/10 km

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004193  
Location = DEVL                    8164.13.00 06/09  
05/12/2010                            13:37

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            659358  
Standard Cyl #?            20  
Standard Expiration?       09/03/2011  
Oper No?                    107501

Flow Cal. Date:            06/04/2009  
Slope                        692  
Intercept                   -611138

IR Calibration Date:        06/03/2009  
                                    3um                    9um

-----  
0th Coef(\*100):            -12132                    -15777  
1st Coef(\*100):            261160                    129907  
2nd Coef(\*100):            1958                      971  
H2O adj(mg/l\*10k):        413                        335

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
\_\_\_\_\_  
Operator Signature  
MARGY PEARSON

Remarks:

Print

Form 106-I8000

80-004193  
P. 73/80 RM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004193  
Location = DEVL      8164.13.00 06/09  
05/12/2010      13:51

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:51
02 Std. Sol.	0.019	13:52
03 Room Air	0.000	13:53
04 Std. Sol.	0.019	13:53
05 Room Air	0.000	13:54
06 Std. Sol.	0.020	13:55
07 Room Air	0.000	13:55


08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000000

County = 36

Oper No. = 107501



Operator Signature  
MARGY PEARSON

Remarks:

*Low AC 0.021 AC*

Form 106-I8000

*80-004193  
P. 7<sup>1</sup>/<sub>80</sub> KA*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004193  
Location = DEVL      8164.13.00 06/09  
05/12/2010      13:56

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:57
02 Std. Sol.	0.297	13:58
03 Room Air	0.000	13:58
04 Std. Sol.	0.297	13:59
05 Room Air	0.000	13:59
06 Std. Sol.	0.297	14:00
07 Room Air	0.000	14:01


08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000000

County = 36

Oper No. = 107501

  
\_\_\_\_\_  
Operator/Signature  
MARGY PEARSON

Remarks:

*Linearity Cal Check 0.292 AC*

Form 106-I8000

80-004193  
P. 75/80 FM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004193  
Location = DEVL      8164.13.00 06/09  
05/12/2010      13:44

Test	AC	Time
01 Room Air	0.000	13:45
02 *Subject Test	INT*	13:45
03 Room Air	0.000	13:46

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS821456  
Simul Ser No = 00000  
Std Sol No = 000000  
County = 36      Oper No. = 107501

  
Operator Signature  
MARGY PEARSON

Remarks:

*Interferent 0.10 AC + Acetone 0.05%*

Form 106-I8000

80-004193  
P. 76/80 KM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004193  
Location = DEVL                    8164.13.00 06/09  
05/12/2010                            13:41

Test	AC	Time
01 Diagnostic	OK	13:42
02 Room Air	RFI*	13:42
03 Room Air	0.000	13:42

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS821456  
Lot No = 659358  
Cyl No = 20  
Expiration Date = 09/03/2011  
County = 36                            Oper No. = 107501

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

  
\_\_\_\_\_  
Operator Signature  
MARGY PEARSON

Remarks:

*RFI*

Form 106-I8000

80-004193  
P. 77/80 PM


\ Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004193  
Location = DEVL                    8164.13.00 06/09  
05/12/2010                            14:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:07
02 Std. Gas	0.081	14:07
03 Room Air	0.000	14:08
04 Std. Gas	0.081	14:08
05 Room Air	0.000	14:09
06 Std. Gas	0.081	14:09
07 Room Air	0.000	14:09

Lot No = 659358  
Cyl No = 20  
Exp Date = 09/03/2011  
County = 36                            Oper No. = 107501



Operator Signature  
MARGY PEARSON

Remarks:

*Cal Check 0.080 AC*

Form 106-18000

80-004193  
P. 78/80 *HA*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004193  
Location = DEVL      8164.13.00 06/09  
05/12/2010      14:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:10
02 Std. Gas	0.081	14:11
03 Room Air	0.000	14:11
04 Std. Gas	0.081	14:12
05 Room Air	0.000	14:12
06 Std. Gas	0.081	14:13
07 Room Air	0.000	14:13

Lot No = 659358  
Cyl No = 20  
Exp Date = 09/03/2011  
County = 36      Oper No. = 107501

  
\_\_\_\_\_  
Operator Signature  
MARGY PEARSON

Remarks:

*Cal Check 0.080 AC*

Form 106-I8000

*80-004193  
P. 79/80 JHR*




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004193  
Location = DEVL      8164.13.00 06/09  
05/12/2010      14:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:14
02 Std. Gas	0.081	14:14
03 Room Air	0.000	14:15
04 Std. Gas	0.081	14:15
05 Room Air	0.000	14:16
06 Std. Gas	0.082	14:16
07 Room Air	0.000	14:17

Lot No = 659358  
Cyl No = 20  
Exp Date = 09/03/2011  
County = 36      Oper No. = 107501

  
\_\_\_\_\_  
Operator Signature  
MARGY PEARSON

Remarks:

Cal Check 0.080 AC

Form 106-I8000

80-004193  
P. 80/80 KM