



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Sean Murphy</u>		
Location: <u>Rattineau Co. Sheriff's Dept.</u>		
Intoxilyzer® Serial Number: <u>80-003065</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>55</u>	Expiration Date: <u>9-3-11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: <u>0.082</u> % Test 2: <u>0.082</u> % Test 3: <u>0.080</u> %		
(ATTACH TEST RECORD)		

*Sean Murphy*  
 Chemical Test Operator Signature

5-12-11  
 Date

Deb Kashner  
 Reviewed By

20 May 2011  
 Date

80-003065  
 p. 94 of 95  
 xam

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003065  
Location = BOTT      8164.13.00 06/09  
05/12/2011      12:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:39
02 Std. Gas	0.082	12:39
03 Room Air	0.000	12:39
04 Std. Gas	0.082	12:40
05 Room Air	0.000	12:40
06 Std. Gas	0.080	12:41
07 Room Air	0.000	12:41

Lot No = 659358

Cyl No = 55

Exp Date = 09/03/2011

County = 05

Oper No. = 039117



Operator Signature  
SEAN J. MURPHY

Remarks:

Form 106-I8000

80-003065  
p. 95 of 95  
JAM