



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Eugene M. Moberg</u>		
Location: <u>Adams Co So HETT</u>		
Intoxilyzer® Serial Number: <u>80-004204</u>		
Gas Lot Number: <u>15470086A1</u>	Gas Cylinder Number: <u>33</u>	Expiration Date: <u>7-1-12</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: <u>0.079</u> % Test 2: <u>0.078</u> % Test 3: <u>0.079</u> %		
(ATTACH TEST RECORD)		

Eugene M. Moberg  
Chemical Test Operator Signature

5-12-2011  
Date

Deb Kashur  
Reviewed By

23 May 2011  
Date

80-004204  
P102 of 104  
FM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004204  
Location = HETT                    8164.13.00 06/09  
05/12/2011                            14:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:59
02 Std. Gas	0.079	14:59
03 Room Air	0.000	15:00
04 Std. Gas	0.078	15:00
05 Room Air	0.000	15:01
06 Std. Gas	0.079	15:01
07 Room Air	0.000	15:02

Lot No = 15410080A1  
Cyl. No = 33  
Exp Date = 07/01/2012  
County = 01                            Oper No. = 049130

  
Operator Signature  
EUGENE M. MOLBERT

Remarks:

Form 106-I8000

80-004204  
P-103 of 104  
FM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004204  
Location = HETT      8164.13.00 06/09  
05/12/2011      14:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:59
02 Std. Gas	0.079	14:59
03 Room Air	0.000	15:00
04 Std. Gas	0.078	15:00
05 Room Air	0.000	15:01
06 Std. Gas	0.079	15:01
07 Room Air	0.000	15:02

Lot No = 15410080A1  
Cyl No = 33  
Exp Date = 07/01/2012  
County = 01      Oper No. = 049130

---

Operator Signature  
EUGENE M. MOLBERT

Remarks:

Form 106-I8000

80-004204  
P. 104 of 104  
FM