



ETHANOL GAS STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Jonathan Lampi</i>		
Location <i>GFPD</i>	Intoxilyzer® Serial Number <i>80-006489</i>	
Gas Lot Number <i>01316080A2</i>	Gas Cylinder Number <i>026</i>	Gas Expiration Date <i>02-05-2018</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.080* AC
 - B. *0.080* AC
 - C. *0.080* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Jonathan Lampi</i>	Date <i>05-12-16</i> ✓
Reviewed By (Crime Laboratory Use Only) <i>Deb Shanaver</i>	Date <i>16 May 2016</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006489
Location = GFPD 8164.13.00 06/09
05/12/2016 11:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:55
02 Std. Gas	0.080	11:55
03 Room Air	0.000	11:56
04 Std. Gas	0.080	11:56
05 Room Air	0.000	11:57
06 Std. Gas	0.080	11:57
07 Room Air	0.000	11:58

Lot No = 01316080A2
Cyl No = 26
Exp Date = 02/05/2018
County = 18 Oper No. = 039947


Operator Signature
JONATHAN LAMPI

Remarks:

Form 106-I8000

Gas cylinder Install