

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004938

Location: Cavalier CoSO - Langdon

- | | Check When Done |
|---|-----------------|
| A. Is the breath tube heated? | <u>✓</u> |
| B. Is the diagnostic check complete? | <u>✓</u> |
| C. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| D. Print test. (Attach test record.) | <u>✓</u> |
| E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5191</u> Lot No. <u>15160</u> Rep AC <u>0.020</u> | <u>✓</u> |
| F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR 3847</u> Lot No. <u>16060</u> Rep AC <u>0.300</u> | <u>✓</u> |
| G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN <u>DR5190</u> Lot No. <u>ICS4</u> Solution <u>0.10 AC + 0.05% acetone</u> | <u>✓</u> |
| H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) | <u>✓</u> |
| I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) <u>0.080</u> AC | |

Lot No. <u>196/5080A5</u>	Cyl No. <u>15</u>	Exp. Date <u>09/05/2017</u>
Test 1 <u>0.080</u>	Test 4 <u>0.080</u>	Test 7 <u>0.079</u>
Test 2 <u>0.079</u>	Test 5 <u>0.080</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.079</u>	Test 9 <u>0.079</u>
Average <u>0.080</u>		

J. Tank pressure, Level 3 Func D Sub G P: 1006 psi Reg ~1000 psi

K. Remarks/Maintenance record: Annual inspection, replaced o-rings on breath hose & simulator return

10 May 16
Date

12 MAY 2016 ✓
Date

Deb Shanaver
Field Inspector's Signature
Charles E. Edr
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = LANG 8164.13.00 06/09
05/10/2016 16:34

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19615080A5
Standard Cyl #? 15
Standard Expiration? 09/05/2017
Oper No? 777777

Flow Cal. Date: 01/25/2010
Slope 712
Intercept -604165

IR Calibration Date: 04/08/2015
 3um 9um


0th Coef(*100): -24553 -24676
1st Coef(*100): 265364 132923
2nd Coef(*100): 3426 1611
H2O adj(mg/l*10k): 426 442

***** Printer Test End *****



Operator Signature
DEB SHANAVER

Remarks:



Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = LANG 8164.13.00 06/09
05/10/2016 16:47

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:48
02 Std. Sol.	0.017	16:49
03 Room Air	0.000	16:49
04 Std. Sol.	0.017	16:50
05 Room Air	0.000	16:51
06 Std. Sol.	0.018	16:51
07 Room Air	0.000	16:52

08 Sim Temp = 34.0°C

Simul Ser No = DR5191
Std Sol No = 15160
County = 10 Oper No. = 777777


Operator Signature
DEB SHANAVER

Remarks:

low AC 0.020AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = LANG 8164.13.00 06/09
05/10/2016 16:53

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:54
02 Std. Sol.	0.302	16:55
03 Room Air	0.000	16:55
04 Std. Sol.	0.303	16:56
05 Room Air	0.000	16:56
06 Std. Sol.	0.304	16:57
07 Room Air	0.000	16:58

08 Sim Temp = 34.0°C

Simul Ser No = DR3847
Std Sol No = 16060
County = 10 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

High AC 0.300 AC
Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = LANG 8164.13.00 06/09
05/10/2016 16:59

Test	AC	Time
01 Room Air	0.000	16:59
02 *Subject Test	INT*	17:00
03 Room Air	0.000	17:00

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 19615080A5
Cyl No = 15
Expiration Date = 09/05/2017
County = 10 Oper No. = 777777


Operator Signature
DEB SHANAVER

Remarks:

*Interferent detected 0.10AC +
0.05% acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = LANG 8164.13.00 06/09
05/10/2016 17:01

Test	AC	Time
01 Diagnostic	OK	17:01
02 Room Air	RFI*	17:01
03 Room Air	0.000	17:02

*Invalid Test
Inhibited - RFI


Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 19615080A5
Cyl No = 15
Expiration Date = 09/05/2017
County = 10 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB SHANAVER

Remarks:



Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = LANG 8164.13.00 06/09
05/10/2016 16:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:38
02 Std. Gas	0.080	16:38
03 Room Air	0.000	16:39
04 Std. Gas	0.079	16:39
05 Room Air	0.000	16:39
06 Std. Gas	0.080	16:40
07 Room Air	0.000	16:40

Lot No = 19615080A5
Cyl No = 15
Exp Date = 09/05/2017
County = 10 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

Calibration check 0.080AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = LANG 8164.13.00 06/09
05/10/2016 16:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:41
02 Std. Gas	0.080	16:41
03 Room Air	0.000	16:42
04 Std. Gas	0.080	16:42
05 Room Air	0.000	16:43
06 Std. Gas	0.079	16:43
07 Room Air	0.000	16:44

Lot No = 19615080A5
Cyl No = 15
Exp Date = 09/05/2017
County = 10 Oper No. = 777777


Operator Signature
DEB SHANAVER

Remarks:
Calibration check 0.080AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = LANG 8164.13.00 06/09
05/10/2016 16:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:44
02 Std. Gas	0.079	16:45
03 Room Air	0.000	16:45
04 Std. Gas	0.080	16:46
05 Room Air	0.000	16:46
06 Std. Gas	0.079	16:46
07 Room Air	0.000	16:47

Lot No = 19615080A5

Cyl No = 15

Exp Date = 09/05/2017

County = 10

Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

calibration check 0.080AE

Form 106-I8000