



# INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59281 (03-2014)

Serial Number <b>80-006504</b>	Instrument Location <b>Bott</b>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): \_\_\_\_\_
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <b>Tyler Wolfe</b>	Date <b>05-12-16 ✓</b>
Reviewed By (Crime Laboratory Use Only) <b>Deb Shanaver</b>	Date <b>20 May 16</b>

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer           Alcohol Analyzer  
North Dakota Model 8000           SN 80-006504  
Location = BOTT                   8164.13.00 06/09  
05/12/2016                           10:08

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:           Enabled  
Start Test Sequence:       DABACABA  
Display Prelim Rslt?       Yes  
Display Third Digit?       Yes  
Inhib Printer(Y/N)?       No  
Display Volume?           No  
Disable On Memfull?       Yes  
# of Print Copies?         1  
Select Std (D/W/I)?       Dry  
Standard Value?           0.080  
Standard Lot #?           01316080A2  
Standard Cyl #?           29  
Standard Expiration?      02/05/2018  
Oper No?                   060901

Flow Cal. Date:           08/20/2015  
Slope                      681  
Intercept                 -673151

IR Calibration Date:      08/20/2015  
                          3um                   9um

-----  
0th Coef(\*100):       -14464           -14817  
1st Coef(\*100):       255017           131849  
2nd Coef(\*100):       5533             1946  
H2O adj(mg/l\*10k):   700               431

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
TYLER WOLFE

Remarks:


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006504  
Location = BOTT                    8164.13.00 06/09  
05/12/2016                            10:09

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:09
02 Std. Gas	0.079	10:10
03 Room Air	0.000	10:10
04 Std. Gas	0.079	10:11
05 Room Air	0.000	10:11
06 Std. Gas	0.079	10:11
07 Room Air	0.000	10:12

Lot No = 01316080A2  
Cyl No = 29  
Exp Date = 02/05/2018  
County = 05                            Oper No. = 060901

  
\_\_\_\_\_  
Operator Signature  
TYLER WOLFE

Remarks:

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006504  
Location = BOTT      8164.13.00 06/09  
05/12/2016      10:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:38
02 Std. Gas	0.079	10:38
03 Room Air	RFI*	10:38
04 Room Air	0.000	10:39

\*Invalid Test  
Inhibited - RFI

Lot No = 01316080A2  
Cyl No = 29  
Exp Date = 02/05/2018  
County = 05      Oper No. = 060901



Operator Signature  
TYLER WOLFE

Remarks:

Form 106-I8000