

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-002/946

Location: BIA-Belmont

- Check When Done
- A. Is the breath tube heated? ✓
- B. Is the diagnostic check complete? ✓
- C. Is the time, date, and year correct (re-set if necessary)? ✓
- D. Print test. (Attach test record.) ✓
- E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR591 Lot No. 15160 Rep AC 0.020 ✓
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3847 Lot No. 16060 Rep AC 0.300 ✓
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5190 Lot No. ICS4 Solution 0.10AC + 0.05% acetone ✓
- H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 19615080 A5 Cyl No. 20 Exp. Date 09/05/2017

Test 1 0.079 Test 4 0.078 Test 7 0.079

Test 2 0.078 Test 5 0.079 Test 8 0.079

Test 3 0.079 Test 6 0.079 Test 9 0.079

Average 0.079

J. Tank pressure, Level 3 Func D Sub G P: 1105 psi Reg 1000 psi

K. Remarks/Maintenance record: Annual inspection, replaced breath hose & simulator return o-rings

11 May 16
Date

12 May 16 ✓
Date

Deb Shanover
Field Inspector's Signature
Charles Elder
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = BELC 8164.13.00 06/09
05/11/2016 10:03

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:03
02 Std. Sol.	0.017	10:04
03 Room Air	0.000	10:05
04 Std. Sol.	0.018	10:05
05 Room Air	0.000	10:06
06 Std. Sol.	0.018	10:07
07 Room Air	0.000	10:07

08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 15160

County = 63

Oper No. = 777777

Deb Shanaver
Operator Signature
DEB SHANAVER

Remarks:

Low alcohol 0.020AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = BELC 8164.13.00 06/09
05/11/2016 10:09

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:10
02 Std. Sol.	0.301	10:11
03 Room Air	0.000	10:12
04 Std. Sol.	0.300	10:12
05 Room Air	0.000	10:13
06 Std. Sol.	0.299	10:14
07 Room Air	0.000	10:14

08 Sim Temp = 34.0°C

Simul Ser No = DR3847

Std Sol No = 16060

County = 63

Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

High alcohol 0.300AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = BELC 8164.13.00 06/09
05/11/2016 10:18

Test	AC	Time
01 Room Air	0.000	10:19
02 *Subject Test	INT*	10:20
03 Room Air	0.000	10:21

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 19615080A5
Cyl No = 20
Expiration Date = 09/05/2017
County = 63 Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

*Interferent detected 0.16 AC + 0.05%
acetone*

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = BELC 8164.13.00 06/09
05/11/2016 10:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:23
02 Std. Gas	0.079	10:23
03 Room Air	0.000	10:24
04 Std. Gas	0.078	10:24
05 Room Air	0.000	10:25
06 Std. Gas	0.079	10:25
07 Room Air	0.000	10:26

Lot No = 19615080A5
Cyl No = 20
Exp Date = 09/05/2017
County = 63 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = BELC 8164.13.00 06/09
05/11/2016 10:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:26
02 Std. Gas	0.078	10:27
03 Room Air	0.000	10:27
04 Std. Gas	0.079	10:28
05 Room Air	0.000	10:28
06 Std. Gas	0.079	10:28
07 Room Air	0.000	10:29

Lot No = 19615080A5
Cyl No = 20
Exp Date = 09/05/2017
County = 63 Oper No. = 777777

Deb Shanaver
Operator Signature
DEB SHANAVER

Remarks:

Calibration check 0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = BELC 8164.13.00 06/09
05/11/2016 10:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:30
02 Std. Gas	0.079	10:30
03 Room Air	0.000	10:31
04 Std. Gas	0.079	10:31
05 Room Air	0.000	10:31
06 Std. Gas	0.079	10:32
07 Room Air	0.000	10:32

Lot No = 19615080A5
Cyl No = 20
Exp Date = 09/05/2017
County = 63 Oper No. = 777777

Deb Shanaver
Operator Signature
DEB SHANAVER

Remarks:

Calibration check 0.080 AC