

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005357

Location: UND PD

- Check When Done
- A. Is the breath tube heated? ✓
- B. Is the diagnostic check complete? ✓
- C. Is the time, date, and year correct (re-set if necessary)? ✓
- D. Print test. (Attach test record.) ✓
- E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5191 Lot No. 15160 Rep AC 0.020 ✓
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3847 Lot No. 16060 Rep AC 0.300 ✓
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5190 Lot No. ICS4 Solution 0.10AC + 0.05% acetone ✓
- H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 14214080A6 Cyl No. 58 Exp. Date 07/05/2016

Test 1 0.079 Test 4 0.079 Test 7 0.080

Test 2 0.079 Test 5 0.079 Test 8 0.079

Test 3 0.079 Test 6 0.079 Test 9 0.078

Average 0.079

J. Tank pressure, Level 3 Func D Sub G P: 530 psi Reg 2520 psi

K. Remarks/Maintenance record: Annual inspection, replaced breath hose o-ring

10 May 16
Date

12 May 16 ✓
Date

Deb Shanaver
Field Inspector's Signature

Charles E. Edr
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005357
Location = UNDP 8164.13.00 06/09
05/10/2016 09:18

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 14214080A6
Standard Cyl #? 58
Standard Expiration? 07/05/2016
Oper No? 777777

Flow Cal. Date: 08/17/2011
Slope 663
Intercept -589414

IR Calibration Date: 04/07/2015
 3um 9um

0th Coef(*100): -16389 -26777
1st Coef(*100): 269467 141875
2nd Coef(*100): 1088 917
H2O adj(mg/l*10k): 159 330

***** Printer Test End *****



Operator Signature
DEB SHANAVER

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005357
Location = UNDP 8164.13.00 06/09
05/10/2016 09:34

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:35
02 Std. Sol.	0.018	09:36
03 Room Air	0.000	09:36
04 Std. Sol.	0.017	09:37
05 Room Air	0.000	09:37
06 Std. Sol.	0.018	09:38
07 Room Air	0.000	09:39

08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 15160

County = 18

Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Low alcohol 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005357
Location = UNDP 8164.13.00 06/09
05/10/2016 09:43

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:43
02 Std. Sol.	0.294	09:44
03 Room Air	0.000	09:45
04 Std. Sol.	0.295	09:45
05 Room Air	0.000	09:46
06 Std. Sol.	0.295	09:47
07 Room Air	0.000	09:47

08 Sim Temp = 34.0°C

Simul Ser No = DR3847

Std Sol No = 16060

County = 18

Oper No. = 777777

Deb Shanaver

Operator Signature.

DEB SHANAVER

Remarks:

High alcohol 0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005357
Location = UNDP 8164.13.00 06/09
05/10/2016 09:53

Test	AC	Time
01 Room Air	0.000	09:56
02 *Subject Test	INT*	09:57
03 Room Air	0.000	09:58

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 14214080A6
Cyl No = 58
Expiration Date = 07/05/2016
County = 18 Oper No. = 777777


Operator Signature
DEB SHANAVER

Remarks:

Interferent detected 0.10Ac + 0.05% acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005357
Location = UNDP 8164.13.00 06/09
05/10/2016 09:32

Test	AC	Time
01 Diagnostic	OK	09:33
02 Room Air	RFI*	09:33
03 Room Air	0.000	09:34

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 14214080A6
Cyl No = 58
Expiration Date = 07/05/2016
County = 18 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB SHANAVER

Remarks:

RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005357
Location = UNDP 8164.13.00 06/09
05/10/2016 09:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:21
02 Std. Gas	0.079	09:22
03 Room Air	0.000	09:22
04 Std. Gas	0.079	09:22
05 Room Air	0.000	09:23
06 Std. Gas	0.079	09:23
07 Room Air	0.000	09:24

Lot No = 14214080A6
Cyl No = 58
Exp Date = 07/05/2016
County = 18 Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Calibration check 0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005357
Location = UNDP 8164.13.00 06/09
05/10/2016 09:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:25
02 Std. Gas	0.079	09:25
03 Room Air	0.000	09:26
04 Std. Gas	0.079	09:26
05 Room Air	0.000	09:27
06 Std. Gas	0.079	09:27
07 Room Air	0.000	09:28

Lot No = 14214080A6

Cyl No = 58

Exp Date = 07/05/2016

County = 18

Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Calibration check 0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005357
Location = UNDP 8164.13.00 06/09
05/10/2016 09:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:28
02 Std. Gas	0.080	09:29
03 Room Air	0.000	09:29
04 Std. Gas	0.079	09:30
05 Room Air	0.000	09:30
06 Std. Gas	0.078	09:31
07 Room Air	0.000	09:31

Lot No = 14214080A6
Cyl No = 58
Exp Date = 07/05/2016
County = 18 Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Calibration check 0.080 AC

Form 106-I8000