

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005953

Location: TOX2

- |   | Check When Done |
|---|-----------------|
| A. Is the breath tube heated?   | <u>✓</u>        |
| B. Is the diagnostic check complete?  | <u>✓</u>        |
| C. Is the time, date, and year correct (re-set if necessary)?   | <u>✓</u>        |
| D. Print test. (Attach test record.)  | <u>✓</u>        |
| E. Low AC. Use $\leq 0.03$ AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3374</u> Lot No. <u>201501B</u> Rep AC <u>0.020</u>   | <u>✓</u>        |
| F. High AC. Use $\geq 0.25$ AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR 3841</u> Lot No. <u>14270</u> Rep AC <u>0.300</u>   | <u>✓</u>        |
| G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR5144</u> Lot No. <u>1CS2</u> Solution <u>0.10 AC + 0.05% acetone</u> | <u>✓</u>        |
| H. RFI check. Run CMS mode. Key radio during CMS test.<br>(Attach test record.)   | <u>✓</u>        |
| I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.) <u>0.080</u> AC   |                 |

Lot No. <u>29614080A2</u>	Cyl No. <u>4</u>	Exp. Date <u>12/05/2016</u>
Test 1 <u>0.081</u>	Test 4 <u>0.082</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.082</u>	Test 8 <u>0.082</u>
Test 3 <u>0.081</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>
Average <u>0.081</u>		

J. Tank pressure, Level 3 Func D Sub G P: 725 psi Reg 750 psi

K. Remarks/Maintenance record: Annual inspection

13 May 15  
Date

13 MAY 15  
Date

Debi Kashur  
Field Inspector's Signature  
Charles E. Ehr  
Reviewed by



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005953  
Location = TOXL      8164.13.00 06/09  
05/13/2015      10:25

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:26
02 Std. Sol.	0.019	10:27
03 Room Air	0.000	10:27
04 Std. Sol.	0.018	10:28
05 Room Air	0.000	10:29
06 Std. Sol.	0.019	10:29
07 Room Air	0.000	10:30

08 Sim Temp = 34.0°C

Simul Ser No = DR3374  
Std Sol No = 201501B  
County = 08      Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

low AC      0.020 AC

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005953  
Location = TOXL      8164.13.00 06/09  
05/13/2015      10:31

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:32
02 Std. Sol.	0.291	10:33
03 Room Air	0.000	10:33
04 Std. Sol.	0.294	10:34
05 Room Air	0.000	10:35
06 Std. Sol.	0.294	10:35
07 Room Air	0.000	10:36

08 Sim Temp = 34.0°C

Simul Ser No = DR3841

Std Sol No = 14270

County = 08

Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*High AC 0.300 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005953  
Location = TOXL      8164.13.00 06/09  
05/13/2015      10:38

Test	AC	Time
01 Room Air	0.000	10:39
02 *Subject Test	INT*	10:39
03 Room Air	0.000	10:40

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 4  
Expiration Date = 12/05/2016  
County = 08      Oper No. = 777777

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check 0.10AE + 0.05% acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005953  
Location = TOXL                    8164.13.00 06/09  
05/13/2015                            10:43

Test	AC	Time
01 Diagnostic	OK	10:43
02 Room Air	RFI*	10:43
03 Room Air	0.000	10:44

\*Invalid Test  
  Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 4  
Expiration Date = 12/05/2016  
County = 08                            Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005953  
Location = TOXL      8164.13.00 06/09  
05/13/2015      10:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:07
02 Std. Gas	0.081	10:08
03 Room Air	0.000	10:08
04 Std. Gas	0.081	10:09
05 Room Air	0.000	10:09
06 Std. Gas	0.081	10:10
07 Room Air	0.000	10:10

Lot No = 29614080A2  
Cyl No = 4  
Exp Date = 12/05/2016  
County = 08      Oper No. = 777777

*Deb Kashur*  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:  
*Calibration check      0.080 AC*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005953  
Location = TOXL      8164.13.00 06/09  
05/13/2015      10:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:11
02 Std. Gas	0.082	10:12
03 Room Air	0.000	10:12
04 Std. Gas	0.082	10:12
05 Room Air	0.000	10:13
06 Std. Gas	0.081	10:13
07 Room Air	0.000	10:14

Lot No = 29614080A2  
Cyl No = 4  
Exp Date = 12/05/2016  
County = 08      Oper No. = 777777

  
Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080AC*

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005953  
Location = TOXL      8164.13.00 06/09  
05/13/2015      10:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:20
02 Std. Gas	0.081	10:20
03 Room Air	0.000	10:21
04 Std. Gas	0.082	10:21
05 Room Air	0.000	10:22
06 Std. Gas	0.081	10:22
07 Room Air	0.000	10:23

Lot No = 29614080A2

Cyl No = 4

Exp Date = 12/05/2016

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Celebration check 0.080 AC*

Form 106-I8000