

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004947

Location: TOXL

- |  | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes?   | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?  | <u>✓</u>        |
| C. Is the breath tube heated?  | <u>✓</u>        |
| D. Is the diagnostic check complete?   | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?  | <u>✓</u>        |
| F. Print test. (Attach test record.)   | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5144</u> Lot No. <u>    </u> Rep AC <u>0.021</u>   | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5114</u> Lot No. <u>    </u> Rep AC <u>0.298</u>                                | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR5186</u> Lot No. <u>    </u> Rep AC <u>0.100</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)  | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)  |                 |

Lot No. <u>668180</u>	Cyl No. <u>94</u>	Exp. Date <u>21 Sept 2011</u>
Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.080</u>	Test 8 <u>0.081</u>
Test 3 <u>0.081</u>	Test 6 <u>0.081</u>	Test 9 <u>0.082</u>
Average <u>0.081</u>		

L. Tank pressure, Level 3 Func D Sub G P: 357 psi Reg 350 psi

M. Remarks/Maintenance record: Annual inspection

14 May 2010  
Date

Deb Kasher  
Field Inspector's Signature

15 May 2010  
Date

Mary Pearson  
Reviewed by

80-004947  
P. 4/14 PM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004947  
Location = TOXL      8164.13.00 06/09  
05/14/2010      13:59

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-|=|  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On Memfull? Yes  
# of Print Copies? 1  
Select Std (D/W/I)? Dry  
Standard Value? 0.080  
Standard Lot #? 668180  
Standard Cyl #? 94  
Standard Expiration? 09/21/2011  
Oper No? 777777

Flow Cal. Date: 01/19/2010  
Slope 679  
Intercept -681178

IR Calibration Date: 01/19/2010  
3um      9um

	3um	9um
0th Coef(*100):	-16742	-27600
1st Coef(*100):	254292	132484
2nd Coef(*100):	3137	1165
H2O adj(mg/l*10k):	754	569

\*\*\*\*\* Printer Test End \*\*\*\*\*

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

Print Test

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004947  
Location = TOXL      8164.13.00 06/09  
05/14/2010      13:59

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:00
02 Std. Sol.	0.019	14:01
03 Room Air	0.000	14:01
04 Std. Sol.	0.019	14:02
05 Room Air	0.000	14:03
06 Std. Sol.	0.019	14:03
07 Room Air	0.000	14:04

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 0000

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Low AC 0.021AC*

Form 106-18000

*80-004947  
P. 8/14 KR*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004947  
Location = TOXL      8164.13.00 06/09  
05/14/2010      14:04

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:05
02 Std. Sol.	0.294	14:05
03 Room Air	0.000	14:06
04 Std. Sol.	0.295	14:07
05 Room Air	0.000	14:07
06 Std. Sol.	0.295	14:08
07 Room Air	0.000	14:08

08 Sim Temp = 34.0°C

Simul Ser No = 00000  
Std Sol No = 0000  
County = 08      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks: *Linearity test*  
*DK Interference check 0.298AC*

Form 106-I8000

80-004947  
P. 9/14 SEM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004947  
Location = TOXL      8164.13.00 06/09  
05/14/2010      14:10

Test	AC	Time
01 Room Air	0.000	14:10
02 *Subject Test	INT*	14:10
03 Room Air	0.000	14:11

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 668180  
Cyl No = 94  
Expiration Date = 09/21/2011  
County = 08      Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check      0.100 AC + acetone*

Form 106-18000

80-004947  
8-19/14 PM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004947  
Location = TOXL      8164.13.00 06/09  
05/14/2010      14:12

Test	AC	Time
01 Diagnostic	OK	14:13
02 Room Air	RFI*	14:13
03 Room Air	0.000	14:13

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 668180  
Cyl No = 94  
Expiration Date = 09/21/2011  
County = 08      Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

80-004947  
P. 11/14 PM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004947  
Location = TOXL      8164.13.00 06/09  
05/14/2010      13:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:56
02 Std. Gas	0.081	13:56
03 Room Air	0.000	13:57
04 Std. Gas	0.081	13:57
05 Room Air	0.000	13:58
06 Std. Gas	0.081	13:58
07 Room Air	0.000	13:59

Lot No = 668180  
Cyl No = 94  
Exp Date = 09/21/2011  
County = 08      Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.081 AC*

Form 106-18000

80-004947  
P. 12/14 PM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004947  
Location = TOXL      8164.13.00 06/09  
05/14/2010      14:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:14
02 Std. Gas	0.081	14:14
03 Room Air	0.000	14:15
04 Std. Gas	0.080	14:15
05 Room Air	0.000	14:16
06 Std. Gas	0.081	14:16
07 Room Air	0.000	14:17

Lot No = 668180  
Cyl No = 94  
Exp Date = 09/21/2011  
County = 08      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-18000

80-004947  
P. 13/14 KM



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004947  
Location = TOXL      8164.13.00 06/09  
05/14/2010      14:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:18
02 Std. Gas	0.081	14:18
03 Room Air	0.000	14:18
04 Std. Gas	0.081	14:19
05 Room Air	0.000	14:19
06 Std. Gas	0.082	14:20
07 Room Air	0.000	14:20

Lot No = 668180  
Cyl No = 94  
Exp Date = 09/21/2011  
County = 08      Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:  
*Calibration check 0.080 AC*

Form 106-I8000

80-004947  
P. 17/4 KSM